BeneView T8

Patient Monitor

Service Manual

Intellectual Property Statement

SHENZHEN MINDRAY BIO-MEDICAL ELECTRONICS CO., LTD. (hereinafter called Mindray) owns the intellectual property rights to this product and this manual. This manual may refer to information protected by copyrights or patents and does not convey any license under the patent rights of Mindray, nor the rights of others. Mindray does not assume any liability arising out of any infringements of patents or other rights of third parties.

Mindray intends to maintain the contents of this manual as confidential information. Disclosure of the information in this manual in any manner whatsoever without the written permission of Mindray is strictly forbidden. Release, amendment, reproduction, distribution, rent, adaption and translation of this manual in any manner whatsoever without the written permission of Mindray is strictly forbidden.

, MINDRAY and BeneView are the registered trademarks or trademarks owned by Mindray in China and other countries. All other trademarks that appear in this manual are used only for editorial purposes without the intention of improperly using them. They are the property of their respective owners.

Contents of this manual are subject to changes without prior notice.

Revision History

This manual has a revision number. This revision number changes whenever the manual is updated due to software or technical specification change. Contents of this manual are subject to change without prior notice.

- Revision number: 1.1
- Release time: April 2008

© 2006-2008 Shenzhen Mindray Bio-Medical Electronics Co., Ltd. All rights reserved.

FOR YOUR NOTES

Preface

Manual Purpose

This manual provides detailed information about the assembling, dissembling, testing and troubleshooting of the equipment to support effective troubleshooting and repair. It is not intended to be a comprehensive, in-depth explanation of the product architecture or technical implementation. Observance of the manual is a prerequisite for proper equipment maintenance and prevents equipment damage and personnel injury.

This manual is based on the maximum configuration; Therefore, some contents may not apply to your monitor. If you have any question, please contact our Customer Service Department.

Intended Audience

This manual is for biomedical engineers, authorized technicians or service representatives responsible for troubleshooting, repairing and maintaining the patient monitors.

Abbreviations

Abbreviations used in this manual are:

MPM	multi-parameter module
SMR	satellite module rack
CMS	central monitoring system
PCB	printed circuit board

Passwords

A password may be required to access different modes within the monitor. The passwords are listed below:

- User maintenance: 888888
- Factory maintenance: 332888
- Demo mode: 2088

Contents

1 Safety 1-1
1.1 Safety Information 1-1
1.1.1 DANGER 1-2
1.1.2 Warnings 1-2
1.1.3 Cautions 1-2
1.1.4 Notes 1-3
1.2 Equipment Symbols 1-3

2 Theory of Operation 2-1
2.1 Introduction 2-1
2.2 System Connections 2-2
2.2.1 Mounting the Patient Monitor 2-2
2.2.2 Connectors for Peripheral Devices 2-3
2.3 Main Unit 2-4
2.3.1 Input System 2-5
2.3.2 Output System 2-6
2.3.3 Processing and Communications System 2-7
2.3.4 Power Management System 2-9
2.3.5 Equipment Interface System 2-12
2.4 Parameter Module 2-14
2.4.1 Module Infrared Communication Board 2-14
2.4.2 Module Power Board 2-14
2.4.3 Parameter Board 2-14
2.5 Satellite Module Rack 2-15

3 Testing and Maintenance 3-1
3.1 Introduction 3-1
3.1.1 Test Report 3-2
3.1.2 Recommended Frequency 3-2
3.2 Visual Test 3-3
3.3 Power On Test 3-4
3.4 Module Performance Tests 3-4
3.4.1 Mainstream CO_2 Tests 3-4
3.4.2 Sidestream and Microstream CO_2 Module Test 3-5
3.4.3 ECG Tests 3-6
3.4.4 Resp Performance Test 3-7
3.4.5 NIBP Tests 3-8
3.4.6 IBP Tests 3-11

3.4.7 SpO_2 Test 3-13
3.4.8 C.O. Test 3-13
3.4.9 Temp Test 3-13
3.4.10 BIS Test 3-14
3.4.11 RM Test 3-14
3.4.12 ICG Test 3-15
3.4.13 AG Tests 3-15
3.5 Nurse Call Relay Performance Test 3-18
3.6 Analog Output Performance Test 3-18
3.7 Electrical Safety Tests 3-19
3.7.1 Enclosure Leakage Current Test 3-20
3.7.2 Earth Leakage Current Test 3-20
3.7.3 Patient Leakage Current Test 3-21
3.7.4 Patient Auxiliary Leakage Current Test 3-21
3.8 Touchscreen Calibration 3-22
3.9 Recorder Check 3-22
3.10 Factory Maintenance 3-23
3.10.1 Accessing Factory Maintenance Menu 3-23
3.10.2 Drawing Waves 3-23
3.10.3 Recorder 3-23
3.10.4 Software Version 3-24
3.10.5 Monitor Information 3-24
3.10.6 Calibrate NIBP 3-24
3.11 Program Upgrade 3-25

4 Troubleshooting 4-1
4.1 Introduction 4-1
4.2 Part Replacement 4-1
4.3 Patient Monitor Status Check 4-1
4.4 Software Version Check 4-2
4.5 Technical Alarm Check 4-2
4.6 Troubleshooting Guide 4-2
4.6.1 Power On/Off Failures 4-2
4.6.2 Display Failures 4-3
4.6.3 Module Rack Failures 4-4
4.6.4 Alarm Problems 4-6
4.6.5 Button and Knob Failures 4-7
4.6.6 Recorder Failures 4-7
4.6.7 Interface Failures 4-8
4.6.8 CF Card Problems 4-9
4.6.9 Power Supply Failures 4-9
4.6.10 Network Related Problems 4-11
4.6.11 Software Upgrade Problems 4-12

5 Repair and Disassembly 5-1
5.1 Tools 5-1
5.2 Preparations for Disassembly 5-1
5.3 Basic Disassembly 5-2
5.3.1 Disconnecting the Base 5-2
5.3.2 Separating the Front and Rear Half of the Monitor 5-4
5.4 Further Disassembly 5-6
5.4.1 Removing the Power Switch & LED Board 5-6
5.4.2 Disconnecting the Knob Encoder 5-7
5.4.3 Removing the Button Board 5-7
5.4.4 Removing the LCD Assembly 5-8
5.4.5 Removing the Alarm Lamp Board 5-12
5.4.6 Removing the Main Board 5-13
5.4.7 Removing the CF/WLAN Assembly 5-14
5.4.8 Removing the Fan 5-16
5.4.9 Removing the Speaker 5-16
5.4.10 Removing the Interface Board Assembly 5-17
5.4.11 Removing the Power Supply Assembly 5-18
5.4.12 Removing the Integral Module Rack 5-21
5.4.13 Removing the Recorder 5-24
5.4.14 Removing the wireless AP 5-30
5.5 Removing the SMR Assembly 5-31
5.6 Disassembling Modules 5-35
5.6.1 Disassembling the ICG Module 5-35
5.6.2 Disassembling CO_2 Module 5-38

6 Parts 6-1
6.1 Introduction 6-1
6.2 Main Unit 6-2
6.3 Base Assembly 6-3
6.4 Front housing Assembly 6-4
6.4.1 17” LCD with Anti-glare Screen 6-4
6.4.2 17” LCD Touchscreen 6-6
6.5 Rear Housing Assembly 6-8
6.5.1 Rear Housing Assembly 6-8
6.5.2 Power module 6-9
6.5.3 Integral Module Rack 6-11
6.5.4 Interface Board Assembly 6-12
6.5.5 Main Support Assembly 6-13
6.5.6 Others 6-15
6.6 SMR Assembly 6-16
6.6.1 SMR Assembly 6-16
6.6.2 SMR Inside Assembly 6-17
6.7 Parameter Modules 6-18

6.7.1 MPM Module 6-18
6.7.2 C.O. Module 6-19
6.7.3 RM Module 6-21
6.7.4 ICG Module 6-22
6.7.5 AG Module 6-24
6.7.6 BIS Module 6-25
6.7.7 IBP Module 6-27
6.7.8 Mindray CO2 Module 6-28
6.7.9 Oridion CO2 Module 6-30
6.8 Wireless AP 6-32
6.9 Replaceable Parts 6-33
6.9.1 Main Unit 6-33
6.9.2 SMR 6-35
6.9.3 Parameter Modules 6-35
6.9.4 Cables 6-36

1 Safety

1.1 Safety Information

⚠DANGER

- **Indicates an imminent hazard that, if not avoided, will result in death or serious injury.**

⚠WARNING

- **Indicates a potential hazard or unsafe practice that, if not avoided, could result in death or serious injury.**

⚠CAUTION

- **Indicates a potential hazard or unsafe practice that, if not avoided, could result in minor personal injury or product/property damage.**

NOTE

- **Provides application tips or other useful information to ensure that you get the most from your product.**

1.1.1 DANGER

There are no dangers that refer to the product in general. Specific “Danger” statements may be given in the respective sections of this manual.

1.1.2 Warnings

⚠WARNING

- **All installation operations, expansions, changes, modifications and repairs of this product are conducted by Mindray authorized personnel.**
- **There is high voltage inside the equipment. Never disassemble the equipment before it is disconnected from the AC power source.**
- **When you disassemble/reassemble a parameter module, a patient leakage current test must be performed before it is used again for monitoring.**
- **The equipment must be connected to a properly installed power outlet with protective earth contacts only. If the installation does not provide for a protective earth conductor, disconnect it from the power line and operate it on battery power, if possible.**
- **Dispose of the package material, observing the applicable waste control regulations and keeping it out of children’s reach.**

1.1.3 Cautions

⚠CAUTION

- **Make sure that no electromagnetic radiation interferes with the performance of the equipment when preparing to carry out performance tests. Mobile phone, X-ray equipment or MRI devices are a possible source of interference as they may emit higher levels of electromagnetic radiation.**
- **Before connecting the equipment to the power line, check that the voltage and frequency ratings of the power line are the same as those indicated on the equipment’s label or in this manual.**
- **Protect the equipment from damage caused by drop, impact, strong vibration or other mechanical force during servicing.**

1.1.4 Notes

NOTE

- **Refer to Operation Manual for detailed operation and other information.**

1.2 Equipment Symbols

	Attention: Consult accompanying documents (this manual).		CIS connector
	Danger: High-voltage		Network connector
	Alternating current(AC)		Defibrillator connector
	Power ON/OFF		Connector for satellite module rack
	Battery indication		Video output
	Zero key		Auxiliary output connector
	Calibrate key		USB connector
	Measure/Standby		Equipotential terminal
	Check sensor		CE marking
	ESD warning symbol for Electrostatic sensitive devices.		
	Type CF applied part. Defibrillator-proof protection against electric shock.		
	Type BF applied part. Defibrillator-proof protection against electric shock.		

FOR YOUR NOTES

2 Theory of Operation

2.1 Introduction

This patient monitor is designed to monitor a fixed set of physiological parameters including ECG, heart rate (HR), respiration (Resp), temperature (Temp), SpO_2, pulse rate (PR), non-invasive blood pressure (NIBP), invasive blood pressure (IBP), cardiac output (C.O.), carbon dioxide (CO_2), oxygen (O_2), anesthetic gas (AG), impedance cardiograph (ICG), bispectral index (BIS) and respiration mechanics (RM) of single adult, pediatric and neonatal patients.



The patient monitor also:

- Provides audible and visual alarm indications in case of patient or equipment problems.
- Enables displaying, reviewing, storing and transferring of real-time data.
- Incorporates multiple input devices such as buttons, knob, touchscreen, keyboard and mouse.
- Interfaces a clinical information system or central monitoring system.
- Enables program upgrade over the network.

2.2 System Connections

2.2.1 Mounting the Patient Monitor

The patient monitor can be mounted on a wall bracket or on a trolley support. The wall bracket or trolley support can be ordered optionally. Each type of mounting bracket is delivered with a complete set of mounting hardware and instructions. Refer to the documentation delivered with the mounting hardware for instructions on assembling mounts.

⚠CAUTION

- **Use mounting brackets we supply or approve. If other compatible mounting bracket is used, be sure it can be safely used on the patient monitor.**
- **The mounting bracket should be installed by our qualified service personnel, or engineers who have adequate knowledge on it.**
- **If other mounting solution is used, the installation personnel and the customer should verify if it can be safely used on the patient monitor, and the customer assume the responsibility for any risk resulting from that.**

2.2.2 Connectors for Peripheral Devices

On the back of the patient monitor you will find all connectors for peripheral devices.



1. AC Power Connector: used to connect an AC power source (100 to 240 VAC, 50/60Hz).
2. Equipotential Terminal: used to connect the equipotential terminal of other equipment, eliminating potential difference between different pieces of equipment.
3. SMR Connector: It outputs a 12V DC, used to connect the SMR.
4. Video Output: It is a DVI-D connector used to connect a secondary display.
5. CIS Connector: It is a RJ45 connector used to connect a CIS.
6. Network Connector: It is a RJ45 connector used to connect an ethernet network or a PC.
7. Defibrillator Connector: It is a RJ11 connector used to connect a defibrillator.
8. Auxi Output Connector: It is a BNC connector used to output analog signals and nurse call signals.
9. Secondary USB Connector: used to connect the mouse and keyboard of the secondary display.
10. General USB Connector: used to connect any USB-compatible peripheral device.

2.3 Main Unit

The patient monitor consists of:

- Input system: button board, knob, touchscreen, power switch and LED board
- Output system: LCD panel, alarm LED board, recorder, speaker
- Processing and communications system: main board, CIS assembly, integral module rack
- Power management system: battery, battery interface board, power module
- Equipment interface system: USB interface board, DVI interface board and CF/WLAN assembly.

Additionally, the patient monitor can also connect a satellite module rack (SMR), parameter modules, mouse, keyboard, etc.

The following diagram illustrates the structure of the patient monitor.



2.3.1 Input System

Button board

The button board, located at the lower part of the monitor's front panel, contains 6 keys and provides connections for the following components to the main board:

- Knob
- Power switch & LED board
- Touchscreen control board
- Backlight board
- Alarm LED board

The following diagram shows the button board connections.



Knob

The knob can be pressed, or rotated both clockwise and counter-clockwise. It is connected with the button board.

Touchscreen

The touchscreen enables touch operations and can be calibrated. It is connected with the touchscreen control board and main board.

Power switch & LED Board

The power switch & LED board controls the power supply for the main unit. It has three LEDs, which respectively indicate the AC power status, battery status and monitor power on/off status. It is connected with the button board.

2.3.2 Output System

LCD

The patient monitor adopts a high-resolution LCD. The LCD is connected with the main board. The signals and power supply from the backlight board are transferred by the button board.

Alarm Lamp

The patient monitor has two alarm lamps: physiological alarm lamp and technical alarm lamp. Physiological alarm lamp lights either red or yellow whereas technical alarm lamp lights blue only. The alarm lamp signals are transferred by the button board and are directly controlled by the main board

Recorder

The recorder receives data coming form the main board and then sends them to the thermal printhead for printing. The recorder has a hardkey (starting/stopping recordings) and a green LED on its front. It is connected with the main board.

The following diagram shows its operating principle.



Module	Description
Power interface	Introduces a DC from the main board.
Power module	Converts the input power into voltages that fit each module and then forwards them to each module.
CPU	Control the communications between modules.
Signal interfac	Control the communications between the main board and the recorder CPU.
Motor drive circuit	Receives the control signals from the CPU and then forwards them to the step engine.s
Button & LED board	Includes one button and one LED which are directly controlled by the CPU.

Speaker

The speaker provides sound for alarms, key strokes, heart beats and pulse, and allows PITCH TONE and multi-level tone modulation. It is connected with the main board and is directly driven by the main board.

2.3.3 Processing and Communications System

Main Board

The main board is the heart of the patient monitor. It implements a series of tasks including input & output control, data storage and processing, display processing, system control, communication management, printing management and alarming, etc.

The main board comprises the CPU board and mother board. The following diagram shows interfaces to other components.



The CPU board is an essential CPU system containing the CPU, FLASH, memory, realtime clock, EEPROM, etc. It interfaces to the mother board only, which then provides interfaces to all other external devices.

The mother board is in charge of connections and communications with other components and provides the following interfaces:

- LCD port: connects a built-in display.
- Video output+network+ GPIO: connects the digital video interface board.
- USB×2+RS422: connects the USB interface board.
- Button board port: connects the button board.
- Integral module rack port: connects integral module rack communication board.
- Fan port: connects the fan.
- Speaker port: connects the speaker.
- Power module port: connects the power module.
- CIS port: located at the back of the mother board for connecting the CIS components.
- WLAN&CF port: connects the CF/ WLAN assembly.
- Recorder port: connects the recorder.

Integral Module Rack

The patient monitor has two kinds of integral module rack: 2-slot and 5-slot. The integral module rack consists of control board and communication board. The control board is a NIOS II module embedded in the communication board. It implements protocol conversion and infrared communication between the main unit and the parameter modules.

The module rack communication board can be a 2-slot type or a 3-slot type. The 3-slot communication board has an embedded NIOSII control board and communicates the main board directly. The 2-slot communication board is connected with the 3-slot communication board and is controlled by the NIOSII board. Both types of communication board consist of the infrared circuit and module power circuit. The RS422 drive circuit is located on the 3-slot communication board.

2.3.4 Power Management System

Battery

The patient monitor uses two chargeable lithium-ion batteries (11.1 V, 4500 mAh). The battery compartment door is located at the bottom of the patient monitor. The battery power is introduced to the power module via the battery interface board, and then processed and distributed to each component by the power module.

NOTE

- **Two batteries must be used simultaneously when the patient monitor operates on battery power. Otherwise, it may cause power supply protection.**

Battery Interface Board

The battery interface board connects batteries to the power module, enabling charging and discharging between the batteries and the power board.

Power Module

The power module is located at the back of the patient monitor. The main part of the power module is the power board, which contains 4 PCBs: charging & power management board, voltage drop DC inverter, voltage rise and drop DC inverter, and voltage drop 5 V CIS power board.

The power module transforms the input power into DC and then forwards them to each component of the patient monitor. The input power comes from either the batteries or an AC source. The patient monitor will run power from the AC source whenever an AC source is available. If the AC source becomes unavailable, the patient monitor will automatically switch to the battery power. This does not affect the monitor's operating status.

Power module has an AC input socket at its backside, and a socket at its front provides 4 connections to the batteries, main board, CIS components and USB interface board respectively. The power module protects itself and the patient monitor by switching off AC input or DC output in case of overcurrent, short circuit and overvoltage.The power module provides 4 DC outputs:

Outputs	Description
+3.3 V	Goes to the mother board, CPU board, DVI interface board and integral module rack.
+5.0 V	Goes to the LCD, DVI interface board, recorder, CF storage card board and USB interface board.
+5.0 V CIS	Goes to the CIS assembly.
+12 V	Goes to the recorder, LCD inverter, integral module rack, parameter modules and USB interface board.

The systematic principle diagram of the power module is as follows:



The following diagram shows the pins of the power module socket (excluding the pins of the battery power socket. On power board, pin 1 has a triangle symbol):


1 2 3 4 5 6 7 8 9 10 11
12V GND 5VCIS GND 5V GND GND RXD TXD PCON LCD-EN
12V GND 5VCIS GND 5V GND 3.3V 3.3V LEDAC LEDBAT LCD-BR
12 13 14 15 16 17 18 19 20 21 22

Pin ID	Marking	Description	Cable color
2, 4, 6, 7, 13, 15, 17	GND	The output grounding terminal of the power board.	Black
8	RXD	Receives serial communications (the main board sends).	Purple
9	TXD	Sends serial communications (the main board receives).	Brown
10	PCON	Power on/off control signal. It is a TTL pulse signal inputted from the back board. Every time when the power on/off switch is pressed (pulse of falling edge), a switch between power “on” and “off” happens. The pulse duration is no less than 0.1 s for power-on and no less than 2 s for power off.	Blue
11	LCD-EN	Backlight on/off control signal. The main board sends a backlight on/off control signal to the power board through the serial interface. The power board processes the received signal and then outputs a high or low level depending on the received signal.	Green
12, 1	12 V	The positive end of the 12 V DC coming from the power board.	Yellow
14, 3	5 V CIS	The positive end of the 5 V CIS coming from the power board.	Purple
16, 5	5 V	The positive end of the 5 V DC coming from the power board.	Red
18, 19	3.3 V	The positive end of the 3.3 V DC coming from the power board.	Orange
20	LED- AC	AC power status indication signal	White
21	LED- BAT	Battery status indication signal.	Grey
22	LCD-BR	Backlight brightness control voltage.	Brown

2.3.5 Equipment Interface System

USB Interface Board

The USB interface board is compatible with such USB interfaces as USB2.0, USB1.1 and USB1.0. It is connected with the main board and the power module. It receives 2 USB differential signals coming from the main board and then distributes them to 10 USB interfaces via 2 ISP1521 chips. Moreover, it provides RS232 and RS422 interfaces for communication between the main board and the SMR. The USB interface board receives 5 VDC and 12 VDC inputs from the power module, of which the 5 VDC goes to the USB interface board and the 12 VDC outputted to the SMR connector through a fuse.



DVI Interface Board

The DVI interface board is connected with the mother board and the CIS mother board. The following diagram shows its interfaces to other components.



Interface	Description
CIS Connector	Connects the CIS mother board.
BNC connector	Outputs analog signals and nurse call signals.
RJ 11 connector	Outputs defibrillator synchronization signals.
RJ 45 connector (network)	It is a standard RJ45 connector, providing 10/100 BASE-TX Ethernet communications channels. It connects an Ethernet network or a PC.
RJ 45 connector (CIS)	It is a standard RJ45 connector for connecting a CIS network.
DVI-D connector	Connects a secondary display.

WLAN&CF assembly

The CF assembly serves the non-volatile CF storage card which is used for data storage and transfer. It is connected with the mother board.

2.4 Parameter Module

Each parameter module consists of the module infrared communication board, module power board, module button board, parameter board, etc.

2.4.1 Module Infrared Communication Board

The module infrared communication board allows a short delay when powering up the module and adopts FPGA to enable infrared communications between the module and the module rack. An ID is integrated into the module infrared communication board. When a module is inserted in the module rack, the ID is automatically sent to the module rack.

2.4.2 Module Power Board

Some modules have no power board. There are two kinds of module power board:

1. Isolated power board: converts the 12 V DC into a 12 V isolated DC and a 5 V isolated DC.
2. Non-isolated power board: converts the 12 V DC into a 5 V DC.

2.4.3 Parameter Board

The parameter board is a parameter measurement component, which is the most important component of the parameter module.

2.5 Satellite Module Rack

The satellite module rack (SMR) is independent of the patient monitor, provides 8 slots for mounting parameter modules. It has the following features:

- It allows a parameter module to be plugged and unplugged with the patient monitor on. This allows function extension and patient transfer.
- It does not have its own power. It introduces 12 V DC from the patient monitor and then supplies power to each parameter module via the contact screw.
- It accomplishes communications protocol conversions between the patient monitor and each parameter module, provides infrared communications for parameter modules, and is responsible for detecting infrared communications malfunction for each parameter module.

The following diagram shows the structure of the SMR.



FOR YOUR NOTES

3 Testing and Maintenance

3.1 Introduction

To ensure the patient monitor always functions normally, qualified service personnel should perform regular inspection, maintenance and test. This chapter provides a checklist of the testing procedures for the patient monitor with recommended test equipment and frequency. The service personnel should perform the testing and maintenance procedures as required and use appropriate test equipment.

The testing procedures provided in this chapter are intended to verify that the patient monitor meets the performance specifications. If the patient monitor or a module fails to perform as specified in any test, repairs or replacement must be done to correct the problem. If the problem persists, contact our Customer Service Department.

⚠CAUTION

- **All tests should be performed by qualified service personnel only.**
- **Care should be taken to change the settings in [User Maintenance] and [Factory Maintenance] menus to avoid loss of data.**
- **Service personnel should acquaint themselves with the test tools and make sure that test tools and cables are applicable.**

3.1.1 Test Report

After completing the tests, service personnel are required to record test results in this table and report them to Mindray Customer Service Department.

Test Equipment			
Name		Model/PN	Expiry Date
Test Record			
No.	Test Item	Test Site	Test Results
1			
2			
Conclusion			
Pass/Fail:		Tested by:	Date:

3.1.2 Recommended Frequency

Check/Maintenance Item		**Frequency**
Visual test		1. When first installed or reinstalled.
Power on test		1. When first installed or reinstalled. 2. Following any maintenance or the replacement of any main unit parts.
Mainstream CO_2 test	/	1. If user suspects that the measurement is incorrect. 2. Following any repairs or replacement of respective module. 3. At least once every two years.
Sidestream and Microstream CO_2 tests	Leakage test	
	Calibration	
ECG tests	Performance test	
	Calibration	
Resp test	/	
NIBP tests	Accuracy test	
	Leakage test	
	Calibration	

IBP tests	Performance test	
	Pressure calibration	
SpO_2 test	/	
C.O. test	/	
Temp test	/	
BIS test	/	
RM test	/	
ICG test	/	
AG tests	Performance test	1. If user suspects that the measurement is incorrect. 2. Following any repairs or replacement of AG module. 3. At least once every year
	Calibration	
Nurse call relay performance test	/	If user suspects that the nurse call or analog output does not work well.
Analog output performance test	/	
Electrical safety tests	Enclosure leakage current test	1. Following any repair or replacement of the power module. 2. At least once every two years.
	Earth leakage current test	
	Patient leakage current test	
	Patient auxiliary current test	
Touchscreen calibration	/	1. When the touchscreen appears abnormal. 2. After the touchscreen is replaced.
Recorder check	/	Following any repair or replacement of the recorder.

3.2 Visual Test

Inspect the equipment for obvious signs of damage. The test is passed if the equipment has no obvious signs of damage. Follow these guidelines when inspecting the equipment:

- Carefully inspect the case, the display screen and the buttons for physical damage to
- Inspect the SMR and parameter modules for mechanical damage.
- Inspect all external connections for loose connectors, bent pins or frayed cables.
- Inspect all connectors on the equipment for loose connectors or bent pins.
- Make sure that safety labels and data plates on the equipment are clearly legible.

3.3 Power On Test

This test is to verify that the patient monitor can power up correctly. The test is passed if the patient monitor starts up by following this procedure:

1. Insert two batteries in the battery chamber and connect the patient monitor to the AC mains, the AC mains LED and battery LED light.
2. Press the power on/off switch to switch on the patient monitor. The operating status LED lights up, and the technical and physiological alarm lamps light blue and red respectively.
3. After the start-up screens are displayed, the system sounds a beep indicating the self test on alarm sounds is passed. At the same time, the technical alarm lamp turns yellow, and then turns off together with the physiological alarm lamp. This indicates that the self test on alarm lamps is passed.
4. The patient monitor enters the main screen and start-up is finished.

3.4 Module Performance Tests

3.4.1 Mainstream CO_2 Tests

Follow this procedure to perform the test:

1. Connect the mainstream CO_2 module with the sensor and plug the module in the module rack.
2. Enter the [**CO2 Setup**] menu to set [**Operating Mode**] to [**Measure**]. Wait for CO_2 module warmup.
3. Select [**Start Zero Cal.**] in the [**CO2 Setup**] menu. If zero calibration fails, the prompt information [**CO2 Zero Failed**] is displayed. If it passes, the baseline of waveform returns to zero.
4. Set [**Apnea Time]** to 10 s in the [**Adjust CO2 Limits**] menu.
5. Blow to the CO_2 sensor to generate a CO_2 waveform. Hold the sensor for 10 s and the alarm message [**CO2 Apnea]** is displayed on the screen.
6. Disconnect the CO_2 sensor, verify that prompt information [**CO2 Comm Stop**] is displayed.

3.4.2 Sidestream and Microstream CO_2 Module Test

Leakage test

Follow this procedure to perform the test:

1. Plug the module in the module rack.
2. Enter the [**CO2 Setup**] menu to set [**Operating Mode**] to [**Measure**]. Wait for CO_2 module warmup.
3. Block the gas inlet completely.
   - Sidestream: Check that alarm message [**CO2 FilterLine Err**] is displayed on the screen in 3s. Block the gas inlet for another 30s, if the alarm message does not disappear, the module does not leak.
   - Microstream: Check that alarm message [**CO2 Purging**] is displayed on the screen in 3s. Block the gas inlet for another 30s, if alarm message [**CO2 FilterLine Err**] is shown, the module does not leak.

Calibration

Tools required:

- A gas bottle with 5% CO_2 gas
- T-shape connector
- Tubing

Follow this procedure to perform a calibration:

1. Make sure that the CO_2 module has been warmed up or started up.
2. Select [**Main Menu**]→[**Maintenance >>**]→[**User Maintenance** >>]→enter the required password→[**Maintain CO2 >>**].
3. In the [**Maintain CO2**] menu, select [**Zero**].

4. Connect the gas bottle with the tubing using a T-shape connector as shown below. Check the airway and make sure there are no leaks.



5. Vent the tubing to the CO_2 opening the gas valve.
6. In the [**Calibrate CO2**] menu, enter the vented CO_2 concentration in the [**CO2**] field.
7. In the [**Maintain CO2**] menu, the measured CO_2 concentration, barometric pressure, sensor temperature and current flowrate are displayed. After the measured CO_2 concentration becomes stable, select [**Calibrate CO2**] to calibrate the CO_2 module.
8. If the calibration is finished successfully, the message [**Calibration Completed!**] is displayed. If the calibration failed, the message [**Calibration Failed!**] is displayed. Perform another calibration.

3.4.3 ECG Tests

ECG Performance Test

Tool required:

- Patient simulator

1. Connect the patient simulator to the ECG/Resp connector on the module.
2. Set HR output to 80 bpm in the ECG simulator.
3. The displayed HR value should be 80 ± 1 bpm for 3- and 5-lead ECG monitoring, 80 ± 2 bpm for 12-lead ECG monitoring.

ECG Calibration

The ECG signal may be inaccurate due to hardware or software problems. As a result, the ECG wave amplitude becomes greater or smaller. In that case, you need to calibrate the ECG module.

1. Select the ECG parameter window or waveform area→[**Filter**]→[**Diagnostic**].
2. Select [**Main Menu**]→[**Maintenance>>**].
3. Select [**Calibrate ECG**]. A square wave appears on the screen and the message [**ECG Calibrating**] is displayed.
4. Compare the amplitude of the square wave with the wave scale. The difference should be within 5%. You can also print out the square wave and the wave scale if necessary.

3.4.4 Resp Performance Test

Tool required:

- Patient simulator

1. Connect the patient simulator to the Resp connector on the module.
2. Configure the simulator as follows: base impedance line 1500 Ω; delta impedance 0.5 Ω, respiration rate 40 rpm.
3. The displayed Resp value should be 40 ± 2 rpm.

3.4.5 NIBP Tests

NIBP Accuracy Test

Tools required:

- T-shape connector
- Approprating tubing
- Balloon pump
- Metal Vessel with volume 500 ± 25 ml
- Reference manometer with accuracy 1 mmHg



Follow this procedure to perform the test:

1. Connect the equipment as shown above.
2. Before inflation, the reading of the manometer should be 0. If not, disconnect the airway and reconnect it unit the readings is 0.
3. Select [**Main Menu**]→[**Maintenance >>**]→[**NIBP Accuracy Test**].
4. Compare the manometer values with the displayed values. The difference between the manometer and displayed values should be no greater than 3 mmHg.
5. Raise the pressure in the metal vessel to 50 mmHg with the balloon pump. Repeat step 3 and 4.
6. Raise the pressure in the metal vessel to 200 mmHg with the balloon pump. Repeat step 3 and 4.

NOTE

- **You can use an NIBP simulator to replace the balloon pump and the reference manometer to perform the test.**

NIBP Leakage Test

Tools required:

- NIBP cuff for adult patient
- Approprating tubing
- Cylinder

Follow this procedure to perform the test:

1. Set [**Patient Category**] to [**Adult**].
2. Connect the NIBP cuff with the NIBP connector on the moduel.
3. Apply the cuff to the cylinder as shown below.
4. Select [**Main Menu**]→[**Maintenance>>**]→[**NIBP Leakage Test**]. The message [**Leakage Testing…**] is displayed in the NIBP parameter area.
5. The cuff automatically deflates in 20s which means NIBP leakage test is completed.
6. If the system does not leak, no message is displayed in the NIBP parameter area. If it leaks, the message [**NIBP Pneumatic Leak**] is displayed. In this case, check for loose connection and perform the test again.



NIBP Calibration

Tools required:

- T-shape connector
- Approprating tubing
- Balloon pump
- Metal Vessel with volume 500 ± 25 ml
- Reference manometer with accuracy 1 mmHg

1. Connect the equipment as shown below.
2. Before inflation, the reading of the manometer should be 0. If not, disconnect the airway and reconnect it unit the readings is 0.
3. Select [**Main Menu**]→[**Maintenance >>**]→[**Factory Maintenance >>**]→enter the required password→[**Calibrate NIBP >>**].
4. Set [**NIBP Pressure**] to 150 mmHg in the [**NIBP measurement circuit**]. Raise the pump output to 150 mmHg. After the pressure value is stabilized, select the [**Calibrate**] button to start the calibration.
5. Set patient category to [**Adu/Ped**] in the [**overpressure protection circuit**], raise the pressure to 330 mmHg. Then, select [**Calibrate**] to start the calibration.
6. Set the patient category to [**Neo**] in the [**overpressure protection circuit**], raise the pressure to 165 mmHg. Then, select [**Calibrate**] to start the calibration.
7. Calibration result is displayed in the [**Calibrate NIBP**] menu. If calibration fails, Check the system connection and perform calibration test again.

3.4.6 IBP Tests

IBP Performance Test

Tool required:

- Patient simulator

1. Connect the patient simulator to the pressure connector on the module.
2. Set the pressure value of patient simulator to 0.
3. Press the Zero Key on the module to start a zero calibration.
4. Configure the patient simulator as P(static)=200 mmHg.
5. The displayed value should be 200 ± 2 mmHg.
6. If the value is beyond tolerance, calibrate the pressure module.

IBP Pressure Calibration

Tools required:

- Standard sphygmomanometer
- Balloon pump
- Tubing
- T-shape connector

1. Connect the equipment as shown below



2. Zero the transducer. After a successful zero, open the stopcock to the manometer.
3. Select [**Main Menu**]→[**Maintenance** >>]→[**Cal**. **IBP Press.** >>]. In the [**Cal. IBP Press.**] menu, enter calibration values for calibrated pressures.
4. Inflate using the balloon pump until the mercury in the manometer reaches to the preset calibration pressure value.
5. Adjust the preset calibration value until it equals to the reading on the manometer.
6. Select the [**Calibrate**] button to the right of the calibrated pressure. The monitor starts a calibration.
7. If the calibration is completed successfully, the message [**Calibration Completed!**] will be displayed. If not, respective message will be displayed.

3.4.7 SpO_2 Test

Tool Required: Patient simulator.

1. Connect the patient simulator to the SpO_2 connector on the MPM.
2. Select the model and the manufacturer of the SpO_2 module under test, and then configure the patient simulator as follows: SpO_2 96%; PR 80 bpm.
3. The displayed SpO_2 and PR values should be within the ranges listed below.

		SpO_2	PR (bpm)
Mindray		96% ± 2%	80 ± 3
Masimo		96% ± 2%	80 ± 3
Nellcor	MAX-A, MAX-AL, MAX-N, MAX-P, MAX-I, MAX-FAST	96% ± 2%	80 ± 3
	OxiCliq A, OxiCliq N, OxiCliq P, OxiCliq I	96% ± 2.5%	
	D-YS, DS-100A, OXI-A/N, OXI-P/I	96% ± 3%	
	MAX-R, D-YSE, D-YSPD	96% ± 3.5%	

3.4.8 C.O. Test

Tool required: Patient simulator

1. Connect the patient simulator to the C.O. connector on the module.
2. Set the patient simulator to output blood temperature (BT) 37°C. The displayed temperature value should be 37 ± 0.1°C.

3.4.9 Temp Test

Tool required: Resistance box

1. Connect the two pins of any Temp connector on the module to the two ends of the resistance box using 2 wires.
2. Set the resistance box to 1354.9 Ω (corresponding temperature is 37°C). The displayed value should be 37 ± 0.2°C.

3.4.10 BIS Test

1. Connect the BIS sensor with a normal and wide awake person as directed in the Operation Manual.
2. The displayed BIS value should be higher than 80.

3.4.11 RM Test

Tool required:

- Gas source
- Ventilator (calibrated)
- Artificial lung
- Flow sensor



Follow this procedure to check the performance of RM module:

1. Connect the equipment as shown above. Make sure that the blue sensing tube on the flow sensor is connected with the artificial lung.
2. Set [**Patient Category**] to [**Adult**]. Enter [**RM Setup**] menu to select [**Sensor Type**] and set [**Ventilation Mode**] to [**Mechanical**].
3. Calibrate the flow sensor.
4. Configure the ventilator or anesthetic machine as follows: TV 500 ml, respiration rate =20 rpm, I:E=1:2.
5. Select [**Respiratory Loop**] in the [**RM Setup**] menu. Verify that displayed value meets product specification.

3.4.12 ICG Test

Tool required: Patient simulator

1. Connect the patient simulator, ICG sensor and the patient monitor using standard ICG cable.
2. Select **[ICG setup]**→[**Patient Demographics** >>]→ input parameter values as follows:

Height	173 cm	CVP	6 mmHg
Weight	75 kg	PAmean	8 mmHg
LV_D	50 mm	Art Mean	93 mmHg
PAWP	10 mmHg		

3. Switch the patient simulator to Mode 1 and then start ICG measurements.
4. After the measurement becomes stable, document the measured values.
5. Switch the patient simulator to mode 2 and then start ICG measurements. After the measurement becomes stable, document the measured values.
6. Check that measured values meet the simulator configuration

3.4.13 AG Tests

AG Performance Test

Tool required:

- Gas bottle with a certain standard anesthetic gas. Gas concentration should meet AA>1.5%, of which AA represents an anesthetic agent
- T-shape connector
- Tubing

1. Plug the AG module in the module rack. Enter [**AG Setup**] menu and set [**Operating Mode**] to [**Measure**].
2. After the AG module warmup is ready, block the gas inlet of the AG module. Alarm message [AC Airway Occluded] is displayed.
3. Connect the equipment as shown below.
4. Open and close the gas valve repeatedly at an interval of 3 seconds. The AA value displayed should meet product specification.

BIS Performance Test

The AG module can incorporates the features of the BIS module, refer to ***3.4.10 BIS Test*** for test procedures.

O_2 Performance Test

Tool required:

- Gas bottle with 100% O_2
- T-shape connector
- Tubing

The AG module can incorporates the features of the O_2 module, refer to ***AG Performance Test*** above for test procedures.

AG Calibration

Tools required:

- Gas bottle, with a certain standard gas or mixture gas. Gas concentration should meet the following requirements: AA>1.5%, CO_2>1.5%, N_2O>40%, O_2>40%, of which AA represents an anesthetic agent.
- T-shape connector
- Tubing
- Reservoir bag

Follow this procedure to perform the calibration:

1. Select [**Main Menu**]→[**Maintenance >>**]→[**User Maintenance** >>]→enter the required password→[**Calibrate AG >>**].
2. Check the airway and make sure that there are no occlusions or leaks.
   - ◆ Vent the sampling tubing to the air and check if the [**Current FlowRate**] and [**Set FlowRate**] are approximately the same. If the deviation is great, it indicates that there is an occlusion in the tubing. Check the tubing for an occlusion.
   - ◆ Block the gas inlet of the tubing. The [**Current FlowRate**] shall fall rapidly and the system prompt that the tubing is blocked. Otherwise, it indicates that there are leakages in the tubing. Check the tubing for leakages.

3. Connect the gas bottle, reservoir bag and the tubing using a T-shape connector as shown in the figure below.



4. Open the gas valve and vent a certain standard gas or mixture gas.
5. In the [**Calibrate AG**] menu, the concentration and flowrate of each measured gas are displayed.
   - If the difference between the measured gas concentration and the actual one is tolerable, a calibration is not needed.
   - If the difference is great, a calibration should be performed. Select [**Calibrate** >>] to enter the calibrate menu.
6. Enter the vented gas concentration. If you use only one gas for calibration, set other gases' concentration to 0.
7. Select [**Start**] to start calibration.
8. If the calibration is finished successfully, the message [**Calibration Completed!**] is displayed. If the calibration failed, the message [**Calibration Failed!**] is displayed. Perform another calibration.

3.5 Nurse Call Relay Performance Test

Tools required:

- Multimeter

1. Connect the nurse call cable to the Auxiliary Output Connector of the patient monitor.
2. Enter Demo mode. Select [**Main Menu**]→[**Maintenance** >>]→[**User Maintenance** >>] →enter the required password→[**Others** >>]→[**Auxiliary Output**]→[**Nurse Call].**
3. Select [**Main Menu**]→[**Maintenance** >>]→[**User Maintenance** >>]→enter the required password→[**Others** >>]→[**Nurse Call Setup** >>]. Enable all alarm level and alarm type options. Set **[Contact Type]** to **[Normally Open]**
4. In [**Nurse Call Setup** >>] setup menu, set [**Signal Type**] to [**Pulse**]. Trigger an alarm and measure the output contact with the multimeter. The output should be square waves with an interval of 1s.
5. In [**Nurse Call Setup** >>] setup menu, set [**Signal Type**] to [**Continuous**]. Trigger an alarm and measure the contact output with the multimeter. The output should be continuous high level.

3.6 Analog Output Performance Test

Tool required:

- Oscillograph

1. Connect the oscillograph to the Auxiliary Output Connector of the patient monitor.
2. Enter Demo mode. Select [**Main Menu**]→[**Maintenance** >>]→[**User Maintenance** >>] →enter the required password→[**Others** >>]→[**Auxiliary Output**]→[**Analog Output].**
3. Select [**Main Menu**] →[**Analog Output Setup].** Switch Analog Output [On], and then test the ECG and IBP signals from the Analog Output Connector.
4. The waveforms displayed on the oscillograph should closely resemble those displayed on the patient monitor.

3.7 Electrical Safety Tests

⚠WARNING

- **Electrical safety tests are a proven means of verifying the electrical safety of the patient monitor. They are intended for determining potential electrical hazards. Failure to find out these hazards timely may cause personnel injury.**
- **Commercially available test equipment such as safety analyzer, etc. can be used for electrical safety tests. Be sure that the test equipment can be safely and reliably used with the patient monitor before use. The service personnel should acquaint themselves with the use of the test equipment.**
- **Electrical safety tests should meet the requirements of the new edition of standards EN 60601-1 and UL60601.**
- **These electrical safety tests do not supersede local requirements.**
- **All devices using the AC mains and connected to medical equipment within patient environments must meet the requirements of the IEC 60601-1 medical electrical systems standard and should be put under electrical safety tests at the frequency recommended for the patient monitor.**

Electrical safety tests are intended to check if potential electrical hazards to the patient, operator or service personnel of the equipment. Electrical safety test should be performed under normal ambient conditions of temperature, humidity and pressure.
The electrical safety test plan described hereafter takes 601 safety analyzer as an example. Different safety analyzers may be used in different areas. Be sure to choose applicable test plan.

Connection of the equipment is shown below:



Tools required:

- Safety analyzer
- Isolation transformer

3.7.1 Enclosure Leakage Current Test

1. Connect the 601 safety analyzer to an AC power supply (264 V, 60 Hz).
2. Connect SUM terminal of the applied part connection apparatus to RA input terminal of 601 safety analyzer, another terminal to the applied part of EUT.
3. Connect the EUT to the analyzer's auxiliary output connector using a power cord.
4. Attach one end of the red lead to the "Red input terminal" of the analyzer, and the other end to tinsel over the enclosure of the EUT.
5. Power on the 601 safety analyzer and then press the "5－Enclosure leakage" button on the analyzer's panel to enter the enclosure leakage test screen.
6. Under normal condition, the enclosure leakage current should be no greater than 100 μA. Under single fault condition, it should be no greater than 300 μA.

3.7.2 Earth Leakage Current Test

1. Connect the 601 safety analyzer to an AC power supply (264 V, 60 Hz).
2. Connect SUM terminal of the applied part connection apparatus to RA input terminal of 601 safety analyzer, another terminal to the applied part of EUT.
3. Connect the EUT to the analyzer's auxiliary output connector using a power cord.
4. Power on the 601 safety analyzer and then press the "4－Earth leakage" button on the analyzer's panel to enter the earth leakage test screen.
5. Under normal condition, the earth leakage current should be no greater than 300 μA. Under single fault condition, it should be no greater than 1000 μA.

3.7.3 Patient Leakage Current Test

1. Connect the 601 safety analyzer to an AC power supply (264 V, 60 Hz).
2. Connect SUM terminal of the applied part connection apparatus to RA input terminal of 601 safety analyzer, another terminal to the applied part of EUT.
3. Connect the EUT to the analyzer's auxiliary output connector using a power cord.
4. Power on the 601 safety analyzer and then press the "6－Patient leakage" button on the analyzer's panel to enter the Patient leakage test screen.
5. Repeatedly press the "Applied Part"button to measure AC and DC leakage alternatively. DC leakage reading is following by "DC".
6. Under normal status, the patient leakage current should be no greater than 10 μA. Under single fault condition, it should be no greater than 50 μA.

3.7.4 Patient Auxiliary Leakage Current Test

1. Connect the 601 safety analyzer to an AC source (264 V, 60 Hz).
2. Connect the equipment under test (EUT) to the analyzer's auxiliary output connector using a power cord.
3. Connect the sensors of the applied part to the applied part connection apparatus, whose RA-P terminal is connected to 601 safety analyzer's RA terminal (which is turned on) and SUM terminal to 601 safety analyzer's LA terminal. RA terminal is switched to on.
4. Power on the 601 safety analyzer and then press the "8－Patient Auxiliary Current Test" button on the analyzer's panel to enter the patient auxiliary current test screen.
5. Repeatedly press the "Applied Part"button to measure AC and DC leakage alternatively. DC leakage reading is following by "DC".
6. Under normal status, the patient auxiliary leakage current should be no greater than 10 μA. Under single fault condition, it should be no greater than 50 μA.

3.8 Touchscreen Calibration

1. Select the [**Cal. Screen**] QuickKey or select [**Main Menu**]→[**Maintenance >>**]→[**User Maintenance >>**]→enter the required password→[**Cal. Touchscreen**].
2. The symbol will appear at different positions of the screen.
3. Select, in turn, the central point of the symbol.
4. After the calibration is completed, the message [**Screen Calibration Completed!**] is displayed. Select [**Ok**] to confirm the completion of the calibration.

3.9 Recorder Check

1. Print ECG waveforms. The recorder should print correctly and printout should be clear.
2. Set the recorder to some problems such as out of paper, paper jam, etc. the patient monitor should give corresponding prompt messages. After the problem is removed, the recorder should be able to work correctly.
3. Switch automatic alarm recording for each parameter ON and then set each parameter's limit outside set alarm limits. Corresponding alarm recordings should be triggered when parameter alarms occur.

3.10 Factory Maintenance

3.10.1 Accessing Factory Maintenance Menu

To access the factory maintenance menu, select [**Main Menu**]→[**Maintenance >>**] → [**Factory Maintenance**]and then enter the required password.

The [**Factory Maintenance**] menu is shown below.



3.10.2 Drawing Waves

There are two methods to draw waves: Color and Mono.

- Color: selecting Color will have smoother waveforms.
- Mono: selecting Mono will have a wider viewing angle.

3.10.3 Recorder

To enable/disable the recorder, select [**Recorder**] and toggle between [**On**] and [**Off**].

⚠CAUTION

- **The recorder is disabled if [Recorder] is switched off in the [Factory Maintenance] menu.**

3.10.4 Software Version

Selecting [**Software Version**] will show software version information. The display of [**Software Version**] menu is as follows:



3.10.5 Monitor Information

Selecting [**Monitor Information**] will show the status of the patient monitor. Monitor information is displayed as follows:



3.10.6 Calibrate NIBP

For details, refer to section ***3.4.5 NIBP Tests*** .

3.11 Program Upgrade

You can upgrade the software of the patient monitor by downloading the upgrade software through network. At the same time, you can view the upgrade status log. The upgrade software can directly run on a portable PC or desktop PC. Through the network or by connecting the patient monitor to a PC via a crossover network cable, you can upgrade the following programs:

- Boot program
- System program
- Multilingual library
- BMP files (including screen icons, start-up screens, standby screens)
- General configurations (including passwords, company logo)
- System functional configuration
- FPGA program
- Integral module rack program
- Satellite module rack program
- Parameter module programs: MPM, IBP module, C.O. module, Mindray CO_2 module, RM module

For details, refer to help and instructions for program upgrade.

⚠CAUTION

- **Disconnect the patient monitor from the patient and make sure the important data are saved before upgrade.**
- **Do not shut down or power off the equipment when upgrading the boot program. Otherwise, it may cause the equipment to break down.**
- **Program upgrade should be performed by qualified service personnel only.**

NOTE

- **After upgrading the boot program, re-upgrade the system program and other programs to ensure compatibility.**
- **Make sure the version of the upgrade package is you desired one. If you want to obtain the latest upgrade package, contact Mindray Customer Service Department.**

FOR YOUR NOTES

4 Troubleshooting

4.1 Introduction

In this chapter, patient monitor problems are listed along with possible causes and recommended corrective actions. Refer to the tables to check the patient monitor, identify and eliminate the troubles.

The troubles we list here are frequently arisen difficulties and the actions we recommend can correct most problems, but not all of them. For more information on troubleshooting, contact our Customer Service Department.

4.2 Part Replacement

Printed circuit boards (PCBs), major parts and components in the patient monitor are replaceable. Once you isolate a PCB you suspect defective, follow the instructions in ***5 Repair and Disassembly*** to replace the PCB with a known good one and check that the trouble disappears or the patient monitor passes all performance tests. If the trouble remains, exchange the replacement PCB with the original suspicious PCB and continue troubleshooting as directed in this chapter. Defective PCB can be sent to us for repair.

To obtain information on replacement parts or order them, refer to ***6 Parts*** .

4.3 Patient Monitor Status Check

Some troubleshooting tasks may require you to identify the hardware version and status of your patient monitor.

1. To view the information on system start time, self check, etc., select [**Main Menu**]→[**Maintenance >>**]→[**Monitor Information >>**].
2. You can also view the information on the monitor's current status by selecting [**Main Menu**]→[**Maintenance>>**]→[**Factory Maintenance>>**]→enter the required password →[**Monitor Information >>**].

4.4 Software Version Check

Some troubleshooting tasks may require you to identify the configuration and software version of your patient monitor.

1. To view information on the system configuration and system software version, Select [**Main Menu**]→[**Maintenance>>**]→[**Software Version>>**].
2. You can also view the information on system software version and module software version by selecting [**Main Menu**]→[**Maintenance>>**]→[**Factory Maintenance>>**]→ enter the required password →[**Software Version>>**].

4.5 Technical Alarm Check

Before troubleshooting the patient monitor, check for technical alarm message. If an alarm message is presented, eliminate the technical alarm first. For detailed information on technical alarm message, possible cause and corrective action, refer to the patient monitor's Operation Manual.

4.6 Troubleshooting Guide

4.6.1 Power On/Off Failures

Symptoms	Possible Cause	Corrective Action
The patient monitor fails to start. AC LED or battery LED does not light	AC mains not connected or battery too low	Check that AC mains is properly connected or battery capacity is sufficient.
	Power supply protection	Refer to ***4.6.9 Power Supply Failures*** .
	Cables defective or poorly connected	1. Check that the cables from power switch & LED board to button board, button board to main board, and power module to main board are correctly connected. 2. Check that cables and connectors are not damaged.
	Power switch & LED board defective	Replace the power switch & LED board.
	Power module defective	Replace the power module.
	Mother board Defective	Replace the mother board.

4.6.2 Display Failures

Symptoms	Possible Cause	Corrective Action
Integrated display is blank but the patient monitor still works correctly.	Cables defective or poorly connected.	1. Check that cables from the display to the mother board and from the backlight board to the button board/display are correctly connected. 2. Check that the cables and connectors are not damaged.
	Backlight board defective	Replace the backlight board.
	Power module defective	Replace the power module.
	Display defective	Replace the display.
Secondary display does not function.	Cables defective or poorly connected.	1. Check that the cable between the display and the patient monitor is correctly connected. 2. Check that the cables and connectors are not damaged.
	DVI interface board defective	Replace the DVI interface board.
Secondary display displays snows or flashing specks	Cables defective or poorly connected.	1. Check that the cable between the display and the patient monitor is correctly connected. 2. Check that the cables and connectors are not damaged.
	DVI interface board defective	Replace the DVI interface board.
	The mother board is damaged.	Replace the mother board.
Images overlapped or distorted	FPGA error.	Update or upgrade FPGA.
	Cables defective or poorly connected.	1. Check that the cable between the display and mother board is correctly connected. 2. Check that the cables and connectors are not damaged.
Touchscreen does not response	Touchscreen disabled	Check if there is a symbol shown above the [**Measurements**] QuickKey. If yes, press the [**Measurements**] QuickKey for more than 3s to enable the touchscreen.

	Cables defective or poorly connected.	1. Check that the cables from the touchscreen to the touchscreen control board, the touchscreen control board to the button board, and the button board to the mother board are correctly connected. 2. Check that the cables and connectors are properly connected
	Touchscreen control board defective	Replace the touchscreen control board
	Button board defective.	Replace the button board.
	Touchscreen defective.	Replace the touchscreen
	Mother board defective	Replace the mother board
Touch position invalid	Touchscreen not calibrated	Calibrate the touchscreen

4.6.3 Module Rack Failures

Symptoms	Possible Cause	Corrective Action
SMR		
SMR cannot identify parameter modules	Extension Cable defective or poorly connected	1. Check that the cable between SMR and main unit is properly connected 2. Check that the connecting cables and connectors are not damaged. 3. Check that contact screws on SMR are tightly screwed and properly contact the SMR.
	Defective parameter module	Replace the suspicious parameter module with a known good module. Check if the patient monitor identifies the replacement module. If yes, it means that the original one is defective.
	Wrong communication board software revision	Upgrade the program of the module or SMR.
	Module (in some slots) unrecognized	1. Replace the Nios II module. 2. Replace the 8-slot module rack communication board.

	Power supply failure	1. Check if the voltage between two contact screws in any slot reaches 12V DC. If yes and the parameter module functions, the PCB assembly in SRM might fail. 2. If there is no 12 V sent to the SMR, check that the power module output voltage to the USB interface board reaches 12V. If yes, the fuse on the USB interface board might blow. Replace the USB interface board.
	Cable defective or poorly connected	1. Check that the cable between SMR interface board and communication board is properly connected. 2. Check that connecting cables and connectors are not damaged.
	Nios II module loose or failure	1. Check that Nios II module is correctly plug ed 2. If the symptom persists, replace the Nios II module.
	SMR interface board failure	Replace the SMR interface board.
	SMR communication board failure	Replace the SMR communication board.
	USB interface board failure	Replace the USB interface board.
	Mother board failure	Replace the mother board.
Integral module rack		
Integral module rack cannot identify parameter modules	Module failure	Replace parameter module. If a new module is identified, the original one is defective.
	Cable defective or poorly connected	1. Check that the cables from 3-slot module rack communication board to MPM module rack communication board, module rack to mother board are properly connected. 2. Check that connecting cables and connectors are not damaged.
	Wrong communication board software revision	Upgrade the program of the module or Integral module rack.
	Module (in some slots) unrecognized	3. Replace the Nios II module. 4. Replace the corresponding module rack communication board.

	Power supply to integral module rack abnormal	1. Check if voltage between two contact screws in any slot reaches 12VDC. If yes and the parameter module functions, PCB assembly in the SMR might fail. 2. If there is no 12V sent to the integrated module rack, check that power module output voltage to mother board reaches 12V DC. If yes, mother board might fail.
	3-slot or MPM module rack communication board failure	Replace the 3-slot or MPM module rack communication board.
	Nios II module failure	Replace the Nios II module.
	Mother board failure	Replace the mother board.

4.6.4 Alarm Problems

Symptoms	Possible Cause	Corrective Action
The alarm lamp is not light or extinguished but alarm sound is issued	Cable defective or poorly connected	1. Check that cables from alarm LED board to button board and button board to mother board are properly connected. 2. Check that connecting cables and connectors are not damaged.
	Alarm LED board failure	Replace the alarm LED board.
	Button board failure	Replace the button board.
	Mother board failure	Replace the mother board.
No alarm sound is issued but alarm lamp is light	Audio alarm disabled	Select [**Main Menu**]→[**Maintenance >>**]→[**User Maintenance >>**]→enter the required password→[**Alarm Setup >>**], and then in the popup menu, set [**Alm Sound**] to [**On**].
	Cable defective or poorly connected	1. Check that cable between speaker and mother board is properly connected. 2. Check that connecting cables and connectors are not damaged.
	FPGA audio logic error	Upgrade the audio logic part of the FPGA program.
	Speaker failure	Replace the speaker.
	Mother board failure	Replace the mother board.

4.6.5 Button and Knob Failures

Symptoms	Possible Cause	Corrective Action
Buttons do not work	Cable defective or poorly connected	1. Check that cable between button board and mother board is properly connected. 2. Check that connecting cables and connectors are not damaged.
	Button board failure	Replace button board.
Knob does not work	Cable defective or poorly connected	1. Check that cables from knob to button board, and button board to mother board are properly connected 2. Check that connecting cables and connectors are undamaged.
	Knob failure	Replace the knob encoder.
	Button board failure	Replace the button board

4.6.6 Recorder Failures

Symptoms	Possible Cause	Corrective Action
No printout	Recorder module disabled	1. Check if the recorder status LED lights 2. If yes, enable the module in [**Factory Maintenance**] menu. Otherwise, check for other possible causes.
	Paper reversed	Re-install the paper roll.
	Cable defective or poorly connected	1. Check that cable between recorder and mother board is properly connected. 2. Check that connecting cables and connectors are not damaged.
	Recorder power supply failure	Check if the power module outputs 5 V DC and 12V DC correctly.
	Recorder failure	Replace the recorder.
Poor print quality or paper not feeding properly	Paper roll not properly installed	Stop the recorder and re-install the paper roll.
	Print head dirty	1. Check the thermal print head and the paper roller for foreign matter. 2. Clean the thermal print head with an appropriate clean solution.
	Recorder failure	Replace recorder.

4.6.7 Interface Failures

Symptoms	Possible Cause	Corrective Action
No analog signals or nurse call signals are issued	Respective output disabled	1. Select [**Main Menu**]→[**Analog Output Setup**]→set [**Analog Output**] to [**On**]. 2. Select [**Main Menu**]→[**Maintenance**>>]→[**User Maintenance** >>]→enter the required password→[**Others** >>]→[**Auxiliary Output**]→select desired output.
	DVI interface board cable loose	1. Check that cable between DVI interface board and mother board is properly connected. 2. Check that connecting cables and connectors are not damaged.
	DVI interface board failure	Replace the DVI interface board.
	Mother board failure	Replace the mother board.
Device with USB port does not function (Assume that the peripheral devices are good)	Cable defective or poorly connected	1. Check that cable between USB interface board and mother board is properly connected. 2. Check that connecting cables and connectors are not damaged.
	USB interface board failure	Replace the USB interface board.
	Mother board failure	Replace the mother board.

4.6.8 CF Card Problems

Symptoms	Possible Cause	Corrective Action
CF card malfunctions	Wrong CF card or small memory space	Use only SanDisk-manufactured CF storage cards. Those with 1GB memory space are recommended.
	Cable defective or poorly connected	1. Check that the cable between CF card board and mother board is correctly connected. 2. Check that connecting cables and connectors not damaged.
	CF card board failure	Replace the CF card board.
	Mother board failure	Replace the mother board.

4.6.9 Power Supply Failures

Symptoms	Possible Cause	Corrective Action
Different battery voltages	Battery failure	Replace battery.
	Cable defective or poorly connected	1. Check that the cable between battery interface board and power module is correctly connected. 2. Check that cables and connectors are not damaged.
	Power board failure	Replace the power board.
Battery capacity is too low	Battery failure	Replace battery.
	Cable defective or poorly connected	1. Check that the cable is correctly connected. 2. Check that connecting cables and connectors are not damaged.
	Power board failure	Replace the power board.
Battery cannot be recharged	Battery failure	Replace battery and recharge the replacement battery. If the replacement battery can be recharged, the original one fails.

	Cable defective or poorly connected	1. Check that cable between battery interface board and power module is correctly connected. 2. Check that cables and connectors are not damaged.
	Power board failure	Replace power board
No +3.3 V output	Power supply protection Power board failure	1. Turn off the monitor then restart it. 2. If the problem remains, disconnect the AC mains for 5 s and reconnect it, then restart the patient monitor. 3. If the problem still remains, replace power board.
No +5.0 V output		
No +5.0 V CIS output		
No +12 V output		

NOTE

- **When the power module has a failure, it may cause problems to other components, e.g. the monitor suddenly breaks down during start-up, as the power module may have a power supply protection. In this case, troubleshoot the power module per the procedure described in the table above.**
- **Components of the main unit, SMR and parameter modules are powered by the power module. In the event that a component malfunctions, check if the operating voltage is correct. Refer to *2 Theory of Operation* for the operating voltage and measurement points of each component.**

4.6.10 Network Related Problems

Symptoms	Possible Cause	Corrective Action
The patient monitor cannot be connected to a CIS.	No connection to LAN	1. Check that the cables and connectors are in good condition and that the network is correctly connected. 2. Check that the hub or switch is correctly configured.
	CIS assembly failure	1. Restart the patient monitor, check that a beep is heard and the CIS fan runs normally after the system starts up. 2. After starting, select [**Main Menu**]→ [**Maintenance>>**]→[**User Maintenance>>**] → enter the password required. In the [**User Maintenance**] menu, [**CIS Maintenance**] shall be enabled and switching between normal monitor screen and CIS screen is possible. 3. If step 1 and 2 fail, check the CIS fan assembly for damage and if the power supply to the CIS assembly is correct. If both of them work correctly, replace the CIS assembly.
	DVI interface board failure	Replace DVI interface board.
Frequent dropouts and network disconnects	Incorrect LAN cable connection	Check LAN cable connection. LAN cable shall not be longer than 50 m.
	Incorrect IP address configuration	Check for IP address conflict. Reconfigure IP address.
The patient monitor is connected to a LAN but cannot view other patients under the View Others mode	Incorrect LAN cable connection	Check LAN cable connection. LAN cable shall not be longer than 50m.
	Excessive requests for viewing the patient monitor at the same time	A patient monitor can only be viewed by 4 other patient monitors at the same time under the View Others mode. The excessive view requests system will be ignored.
	Incorrect IP configuration	Check for IP address conflict. Reconfigure IP address.
	CIS assembly failure	Replace CIS assembly.

4.6.11 Software Upgrade Problems

Symptoms	Possible Cause	Corrective Action
Boot file upgrade fails	Power failure or unintended power off during boot file upgrade	Return the CPU board to factory for repair.
Program upgrade fails	Incorrect network connection	1. Check that network connector, not CIS connector, on the patient monitor is used. 2. Make sure that the hub or switch run normally. Check that net twines are of the right type and have been connected correctly.
	Wrong upgrade package has been downloaded	Upgrade package shall be .pkg files. Select package according to system requirement.
	Incorrect IP address configuration	Configure a fixed IP address in range C as specified for the patient monitor. We recommend not to upgrade a program when the patient monitor is connected to a network with multiple PCs.

5 Repair and Disassembly

5.1 Tools

During disassembly and replacing, the following tools may be required:

- Phillips screwdrivers
- Small flat-bladed screwdrivers
- Contact spanner
- Tweezers
- Sharp nose pliers
- Sleeve

5.2 Preparations for Disassembly

Before disassembling the monitor, finish the following preparations:

- Stop monitoring the patient, turn off the monitor and disconnect all the accessories and peripheral devices.
- Disconnect the AC power source and take out both of the batteries.
- Pull off all the modules in the integral module rack. If the SMR is connected, disconnect the SMR from the monitor and then remove all the modules in it.

⚠WARNING

- **Before disassembling the monitor, be sure to eliminate the static charges first. When disassembling the parts labeled with static-sensitive symbols, make sure you are wearing electrostatic discharge protection such as antistatic wristband or gloves to avoid damaging the equipment.**
- **Put the cables or wires in place when reassemble the monitor to avoid short circuit.**
- **When assembling the monitor, be sure to select proper screws. If an unfit screw is tightened by force, the monitor may be damaged and the screw or the part may fall off during use to cause unpredictable damage or human injury.**

- **Be sure to follow the correct sequence to disassembly the monitor. Otherwise, the monitor may be damaged permanently.**
- **Be sure to disconnect all the cables before disassembling any parts. Be sure not to damage any cables or connectors.**
- **Be sure to place the monitor face up when disassembling it. Otherwise, the screen or the knob may be scratched or damaged.**

5.3 Basic Disassembly

5.3.1 Disconnecting the Base

NOTE

- **Be sure to disassemble the base first before proceeding with other parts.**
- **Be sure to place the monitor face up when disassembling it. If lay the monitor face down, be sure the surface is non-abrasive and static-free. Otherwise, the screen or the knob may be scratched or damaged.**

1. Place the monitor face up and unscrew the four M4×12 screws, as shown in the figure below.

2. Pull out the base and then unplug the two cables marked in the picture, one connecting the Power Switch & LED board and the button board, and the other connecting the battery interface board and the power supply assembly.



NOTE

- **Exercise care when pulling the base out. Be sure not to damage the cables and connectors.**

5.3.2 Separating the Front and Rear Half of the Monitor

1. Keep the monitor (without the base) face up. Disconnect the cables marked in the picture, one connecting the LCD panel and the mother board, and the other connecting the button board and the mother board.



2. Carefully place the monitor face down. Pry up the four screw covers with a small flat-bladed screwdriver and unscrew the four M3×12 screws exposed. Then unscrew the M3×12 screw in the handle.



NOTE

- **Press the cover with a thumb when prying it.**

3. Lift the rear cover assembly to separate it from the front cover assembly.

5.4 Further Disassembly

5.4.1 Removing the Power Switch & LED Board

1. Locate the cable marked in the picture and disconnect it from the power switch & LED board.



2. Release the clips with fingers. At the same time pinch the power switch & LED board, push it to the left and take it out along with the power switch.

NOTE

- **Exercise care whe releasing the clips.**
- **When installing the LED board along with the power switch, put it on the left clip, unclench the right clip to the right direction slightly, and then press it down.**

5.4.2 Disconnecting the Knob Encoder

1. Disconnect the cable that connects the knob encoder and the button board. Then unscrew the two M3×6 screws and take out the knob encoder.



5.4.3 Removing the Button Board

1. Locate the cables marked in the picture and disconnect them from the button board.



2. Unscrew the three PT3×8 screws and take out the button board.

5.4.4 Removing the LCD Assembly

⚠CAUTION

- **Do not touch the LCD screen.**
- **Disassemble the LCD screen in an environment as dust-free as possible.**

1. Disconnect the cables that connect the following parts.
   - ◆ The backlight board and the button board.
   - ◆ The touchscreen control board and the button board.
   - ◆ The alarm lamp board and the button board.
   - ◆ The touchscreen and the touch screen control board



2. Unscrew the two M3×6 screws and take out the screen assembly carefully. If it is necessary to take the touchscreen control board out separately, remove the cables connecting to the button board and touchscreen first.

3. Unscrew the six M3×6 screws. Then remove the cables connecting the backlight board to the LCD screen and to the button board and take out the backlight board by unscrewing the two M3×6 screws on the backlight board.





4. Unscrew the eight M3×6 screws and take the screen assembly out carefully. To prevent the screen from being contaminated by dust, do not touch the screen.



Do not touch the LCD panel



Touchscreen



5.4.5 Removing the Alarm Lamp Board

1. After removing the LCD panel, disconnect the cable that connects the alarm lamp board and the button board, and then remove the alarm lamp board on which there is no screw.

NOTE

- **Exercise care when removing the alarm lamp board because it may be adhered to the LCD assembly (as shown in the upper-right figure).**

5.4.6 Removing the Main Board

1. Pull out all the connectors on the mother board. There are numbers beside the connectors, which are listed below.

Number	Connected to
J1	Power module
J3	Fan
J4	Integral module rack
J5	Button board
J6	DVI interface board
J7	LCD panel
J8	Recorder
J9	WLAN&CF assembly
J10	CIS assembly
J11	USB interface board
J12	Speaker
J22/J23	CPU board, which is connected to the mother board with a socket

2. Unscrew the four M3×6 screws as shown in the figure above and take out the main board that includes the mother board and the CPU board.
3. Unscrew the four M2.5×6 screws and separate the mother board from the CPU board. Be sure not to damage the socket that connects the two boards. There is a battery on the CPU board.



5.4.7 Removing the CF/WLAN Assembly

1. Unscrew the two M3×6 screws.



2. Lift the CF/WLAN assembly and pull it out.



NOTE

- **When reassembling the CF/WLAN assembly, be sure to align it with the WLAN slot. Otherwise, it may be unable to be assembled into position.**

3. Unscrew the four M3×6 screws and separate the CF storage card from the PCMCIA card.





NOTE

- **Be sure not to damage the socket that connects the two boards.**

5.4.8 Removing the Fan

1. Disconnect the connector that connects the fan and the mother board. Then unscrew the three M3×6 screws and remove the fan.



5.4.9 Removing the Speaker

1. Disconnect the connector that connects the speaker and the mother board. Then unscrew the two M3×6 screws and remove the speaker.

⚠CAUTION

- **When installing a speaker, be sure to keep the cable from the press bar. Otherwise, the speaker may be short-circuited.**

5.4.10 Removing the Interface Board Assembly

1. Unscrew the four M3×6 screws and remove the interface board assembly.



2. Unscrew the screws beside the DVI socket. Then unscrew the two M3×6 screws in the holes. After that, pull out the DVI interface board.



Screws beside DVI socket



Two M3×6 Screws in Holes

3. Unscrew the two M3×6 screws and take out the USB interface board upwards.



NOTE

- **Before removing the USB interface board, the DVI interface board must be removed first.**
- **Be careful not to damage the insulation between the DVI interface board and the USB interface board. If it is damaged, stick a new one.**

5.4.11 Removing the Power Supply Assembly

1. Disconnect the cables from the socket of the power supply.



NOTE

- **For some cables, you have to release the clips on it before disconnecting them.**

2. Unscrew the two M4×20 screws. Be sure to hold the power supply assembly to prevent it from falling when unscrewing the screws.



3. Lift the power supply assembly slightly to separate it from the two shafts on the rear cover and then remove the assembly.

Shafts

4. Unscrew the two M3×12 screws and remove the cover of the power supply.

5. Unscrew the two M3×6 screws. pry out the board with a small flat-bladed screwdriver in the gap between the power supply board and the housing. Then turn the board over and unplug the cables from it.





NOTE

- **Because the power supply board may be adhered to the insulator, be careful not to damage the parts, connector and cables when prizing it.**

5.4.12 Removing the Integral Module Rack

The following disassembling procedure takes the 5-slot module rack as an example. You can disassemble other types of module rack by referring to this procedure.

1. Disconnect the cable that connects the integral module rack and the mother board. Remove the two bolts and unscrew the M3×6 screw.



2. Pry the snaps out about 1mm away with a small flat-bladed screwdriver. Then insert the small flat-bladed screwdriver into the position marked in the picture and pry it out about 1mm so that the module rack becomes disengaged from the back cover.

3. If the module rack still engages with the back cover, insert a small flat-bladed screwdriver into the position marked in the picture and pry it out about 1mm to release the hidden snap between the integral module rack and the back cover. Then pull out the module rack.



MPM Communication Board 3-slot module Rack Communication Board

⚠CAUTION

- **Be sure to release the hidden snaps first when removing the integral module rack. Otherwise, the monitor will be damaged.**

4. Release the three clips on the 3-slot module rack communication board and remove the Nios II module.





Nios II module

5.4.13 Removing the Recorder

1. Open the recorder door and unscrew the two M3×6 screws.



2. Pull the two clips in the directions as indicated and meanwhile pull out the recorder.



NOTE

- **Be sure not to damage the connecting cables or connectors when pulling out the recorder.**

3. Unscrew the M3×6 screw and remove the cables marked in the picture.

4. Unclench the two clips and meanwhile take out the recorder drive board. Pay attention to the snap in the front.

5. Pull out the press bar upwards about 1mm and then pull off the flexible cable. Remove the cable that connects the drive board and the button board. Unscrew the PT2×6 screw and remove the drive board's grounding cable. Then take out the recorder drive board.



6. Unscrew the two PT2×6 screws and take out the thermal printhead. Then unscrew the two PT2×6 screws and remove the recorder's button board.



7. Remove the recorder mounting bracket by pulling out the recorder mounting bracket from the right side first.



Recorder mounting bracket

8. Hold and then pinch the recorder housing so that the recorder housing becomes disengaged from the monitor housing. Then take out the recorder housing.



NOTE

- **When pinching the recorder housing, take care not to damage the recorder housing.**
- **When reassembling the recorder housing and the recorder mounting bracket, install the recorder mounting bracket first and meanwhile be sure the two clips are in right position. Otherwise, the recorder cannot be installed correctly.**

5.4.14 Removing the wireless AP

1. Disconnect the network cable and the power cord

2. Press the leaf on the fix board with forceps; push the AP and installation board upwards. Then remove the AP and installation board.



3. Unscrew the two M3×8 screws and take out the fix board.



5.5 Removing the SMR Assembly

1. First remove the 4 screw covers and then unscrew the 4 M3×8 screws.



2. Pull off the left- and right-side boards. Be sure to place the rubber loop in position when reassembling the right board.

3. From the left side, remove the cable that connects the module rack interface board and the 8-slot module rack communication board. Then take off the SMR cover.

4. Unclench the two clips and take out the module rack interface board. Be sure not to damage the snap slot on the left side.



5. Remove the cable that connects the 8-slot communication board and the LED board, the LED indicator and the light tube.



6. Unclench the clips and take out the Nios II module. Then unscrew the six M3×6 screws and remove the 8-slot module rack communication board.



7. Use the sleeve to unscrew the hexagon nut and countersunk external toothed lock washer assembly which can be further separated into the washer, springlet and contact screw.





5.6 Disassembling Modules

⚠WARNING

- **For a reassembled module, a patient leakage current test must be performed before it is used again for patient monitoring.**
- **Make sure the monitor is off before disassembling the modules.**

The monitor supports three types of modules, single-slot modules, 2-slot modules and 3-slot modules. The following part describes two typical disassembling procedures.

5.6.1 Disassembling the ICG Module

1. Unscrew the two contact screws and Φ3 spring washer on the back with a dedicated contact spanner.

2. Move the snap lock in front of the contact spanner to the unlocking position. Insert a small flat-bladed screwdriver into the hole on the snap lock and press the snap down about 1mm. At the same time, push the snap lock forwards with a thumb until the snap lock and its counterpart separate. Then release the snap lock with a screwdriver.



3. Unscrew the M3×6 screw. Then press down, in turn, the two clips that engage the front cover. At the same time, pull off the front cover carefully to avoid damaging the cables.



4. Press the two clips about 1mm and take off the housing.



5. The side cover of the module without the housing is composed of two halves. Release the three snaps to separate the two halves.







6. Take out the non-insulated power board and disconnect all the cables from it.

7. Take out the infrared communication board and disconnect all the cables from it.



8. Disconnect all the cables on the ICG board and take off it with the front cover.





5.6.2 Disassembling CO_2 Module

1. Remove the four contact screws on the back and two snap locks on the bottom and then unscrew the two M3×6 screws by referring to the procedure as described in ***5.6.1 Disassembling the ICG Module*** .

2. Take off the front cover of the CO_2 module by referring to the procedure as described in ***5.6.1 Disassembling the ICG Module***. Then press the two clips about 1mm and take off the housing.





3. Hold the housing with both hands. Press the snap side with the thumb and move the other side with the index finger and the middle finger to separate the housing into two halves.



4. Take out the infrared communication board and disconnect all the cables from it.

5. Disconnect the cables from the button board and unscrew the PT3×8 screw. Then release the two clips and take out the board. After that, disconnect the cable to the fan and remove the button board.





6. Disconnect the tubing that connects the CO_2 board and the front cover. Then release four clips and take off the board.



6 Parts

6.1 Introduction

This section contains the exploded views and parts lists of the main unit, satellite module rack and parameter modules of the patient monitor. It helps the engineer to identify the parts during disassembling the patient monitor and replacing the parts.

Hardware architecture of the main unit is shown below:



6.2 Main Unit

Exploded View



Parts List

SN	P/N	Description	Qty
1	6800-30-50479 or 6800-30-50480	Front housing assembly （17" LCD, anti-glare screen） Front housing assembly（17" LCD, touchscreen）	1
2	M6G-020015---	Hose, 0.47 m	1
3	6800-30-50473	Rear housing assembly	1
4	M04-004017---	Crosshead screw M3×12	5
5	6800-20-50189	Screw cap	4
6	6800-20-50468	Base assembly	1
7	M04-005405---	Screw, M4×12	4

6.3 Base Assembly

Exploded View



Parts List

SN	P/N	Description	Qty
1	6800-20-50183	Battery cover R	1
2	6800-20-50386	Battery cover spacer	2
3	6800-20-50181	Bottom cover	1
4	M04-004012---	Crosshead screw M3×6	6
5	6800-20-50212	Base support	1
6	6800-30-50108	Battery interface board	1
7	6800-20-50201-51	Power switch	1
8	6800-30-50088	Power switch board	1
9	M05-010002-06	Lithium battery, 1.1 V, 4500 mAh	2
10	6800-20-50182	Battery cover L	1
11	6800-20-50233	Cushion	4

6.4 Front housing Assembly

6.4.1 17” LCD with Anti-glare Screen

Exploded View



Parts List

SN	P/N	Description	Qty
1	6800-20-50177	Front housing	1
2	6800-20-50199	Alarm LED cover	1
3	6800-20-50390	Light conduction block 2	1
4	6800-20-50389	Light conduction block 1	1
5	6800-30-50086	Indicator board	1

6	6800-20-50691	Insulating strip for alarm LED board	1
7	6800-20-50220	PC cusion 1	1
8	6800-20-50221	PC cusion 2	1
9	6800-20-50228	Dust-proof pad 7	2
10	6800-20-50227	Dust-proof pad 6	2
11	6800-20-50207	Screen support 4(for 17” anit-glare screen)	1
12	M04-004012---	Crosshead screw M3×6	22
13	6800-20-50232	Backlight board insulating sheet	1
14	M04-002505---	Crosshead screw M3×6	2
15	6800-20-50671	Backlight board shielded cover	1
16	6800-20-50692	Backlight board insulating sheet	2
17	0000-10-11072	Converter DC-AC12VDC/700Vrms	1
18	9901-10-23920	Conducting foam 0501080	2
19	0000-10-10996	Leaf	1
20	6800-20-50203	Screen mounting plate	1
21	M6T-030003---	Conducting foam, 4105AB51K	4
22	0000-10-10754 or 0000-10-11069 (optional)	17”LCD screen	1
23	6800-20-50206	Screen support 3(for 17” anit-glare screen)	1
24	6800-20-50336	Anti-glare screen	1
25	6800-20-50392	PC cusion 4	1
26	6800-20-50391	PC cusion 3	1
27	M04-003105---	Tapping screw PT3×8	3
28	6800-30-50887	keypad (1280x1024-1+anti-glare) (without shank)	1
29	6800-20-50198	Silicon buttons	1
30	6800-30-51132	Knob assembly (without shank)	1

6.4.2 17” LCD Touchscreen

Exploded View



Parts List

SN	P/N	Description	Qty
1	6800-20-50177	Front bezel	1
2	6800-20-50199	Alarm LED cover	1
3	6800-20-50390	Light conducting block2	1
4	6800-20-50389	Light conducting block1	1
5	6800-30-50086	Indicator board	1
6	6800-20-50691	Insulating strip for alarm LED board	1
7	6800-20-50222	Dust-proof pad 1	1
8	6800-20-50223	Dust-proof pad 2	1

9	0000-10-11071 or 0000-10-10748 (optional)	Touchscreen, resistance, 17.1″	1
10	6800-20-50226	Dust-proof pad 5	2
11	6800-20-50205	Screen support 2(for 17” touchscreen)	1
12	6800-20-50232	Backlight board	1
13	M04-004012---	Crosshead screw M3×6	22
14	0000-10-11072	Backlight board	1
15	M04-002505---	Crosshead screw M3×6	2
16	6800-20-50692	Backlight board insulating sheet	2
17	6800-20-50671	Backlight board shield	1
18	0000-10-10996	Berylium-bronze leaf 92-047	1
19	9901-10-23920	Conducting foam0501080	2
20	6800-30-50082 or 6800-30-51095 (optional)	Touchscreen control board （6800-30-51095 only used together with the optional touchscreen 0000-10-10748）	1
21	6800-20-50203	Screen mounting plate	1
22	M6T-030003---	Conducting foam, 4105AB51K	4
23	0000-10-10754 or 0000-10-11069 (optional)	LCD screen	1
24	6800-20-50204	Screen support 1(for 17’ touchscreen)	1
25	6800-20-50225	Dust-proof pad 4	2
26	6800-20-50224	Dust-proof pad 3	1
27	6800-20-50142	Touchscreen waterproof strip	1
28	M04-003105---	Tapping screw PT3×8	3
29	6800-30-50888	keypad (1280x1024-1+Touch)(without shank)	1
30	6800-20-50198	Silicon buttons	1
31	6800-30-51132	Knob assembly (without shank)	1

6.5 Rear Housing Assembly

6.5.1 Rear Housing Assembly

Exploded View



Parts List

SN	P/N	Description	Qty
1	TR6F-30-67306	TR6F recorder	1
2	6800-20-50193	Recorder chamber	1
3	6800-20-50213	Recorder support rack	1
4	6800-20-50196	CF card door	1
5	6800-20-50398	WLAN pushing bar	1
6	6800-20-50185	Handle cover	1
7	6800-20-50184	Rear housing	1

8	M6G-020015---	Hose, 0.3m	1
9	6800-20-50234	Waterproof strip I	1
10	6800-30-50475	Power module assembly	1
11	6800-30-50469	Module rack (maximum configuration)	1
12	6200-21-11629	Speaker	1
13	6800-20-50195	Speaker pad	1
14	M04-004012---	Crosshead screw M3×6	24
15	6800-20-50187	plug	2
16	6800-30-50533	Main support assembly	1
17	6800-30-50476	Interface board assembly	1
18	6800-30-50532	Main board assembly	1
19	6800-30-50472	CF/WLAN assembly	1
20	6800-20-50395	WLAN tray	1
21	M04-051054---	Crosshead screw M4×20	2
22	M04-004015---	Crosshead screw M3×8	2

6.5.2 Power module

Exploded View



Parts List

SN	P/N	Description	Qty
1	0509-20-00098	Grounding terminal	1
2	6800-20-50197	Power module chamber	1
3	6800-20-50692	Backlight board insulating sheet	1
4	6800-20-50693	Backlight board insulating plate I, 100×27 mm	1
5	M04-004017---	Philips pan head screw M3×12	2
6	M04-021024---	Large washer	2
7	6800-20-50219	Sheet metal	1
8	6800-20-50465	Power board insulating sheet	1
9	6800-30-50050	Power board	1
10	M04-004012---	Crosshead screw M3×6	4
11	6800-20-50218	power socket fixture	1
12	6800-20-50636	Waterproof strip for power socket	2
13	6800-21-50337	AC filter connector	1
14	M04-004401---	Stainless steel nut, hexagon GB/T6170-2000 M6 polished	1
15	M04-004504---	Spring washer	1
16	M04-021003---	Flat washer	1

6.5.3 Integral Module Rack

Exploded View



Parts List

SN	P/N	Description	Qty
1	6800-20-50190	Side plate, rear housing-ALL	1
2	6800-20-50191	Side plate small cover, rear housing -ALL	1
3	M04-003105---	Tapping screw PT3×8	2
4	6800-30-51152	Three-slot module rack communication board	1
5	M04-004012---	Philips pan head screw M3×6	9
6	6800-20-51127	Insulating plate	1
7	6800-30-50075	General NiosII module (Module rack Non-lead technique)	1
8	6800-30-50073	MPM module rack communication board	1
9	M04-000301---	Stainless steel hexagon nut GB/T6170-2000 M3 polished	10

10	M04-000802---	Flat washer-grade A GB/T97.1-20023 nickel plated to resist rust	10
11	M04-000104---	Standard spring washerGB/T93-1987 3 nickel plated to resist rust	10
12	6800-20-50388	Leaf	10
13	6800-20-50261	Contact spring	10
14	6800-21-51100	Contact screw	10
15	6800-20-50279	Infrared lens	5

6.5.4 Interface Board Assembly

Exploded View



Parts List

SN	P/N	Description	Qty
1	6800-20-50217	Interface support	1
2	6800-30-50063	DVI interface board	1
3	6800-20-50241	Interface board insulating plate	1
4	6800-30-50065	USB interface board	1
5	6800-20-50745	Waterproof strip	1
6	M04-004012---	Screw, M3×6	4
7	M6T-030005---	Berilium & bronze leaf 187S30（4 leaves）	7
8	M6T-030005---	Berilium & bronze leaf 187S30（2 leaves）	1
9	M6T-030006---	Berilium & bronze leaf 92-106 nickel plated	6

6.5.5 Main Support Assembly

Exploded View



Parts List

SN	P/N	Description	Qty
1	6800-20-50215	Main support	1
2	6800-20-50396	Vent baffle 1	1
3	6800-30-50509	Fan assembly	1
4	M04-051045-00	Stainless steel Philips screw M2.5X4	4
5	0000-10-10755	Control board	1
6	6800-30-50112	CIS mother board	1
4	M04-004012---	Crosshead screw M3×6	6
5	6800-20-50550	CIS hard disk support	1
8	6800-20-50550	CIS hard disk support rack	1
9	0000-10-11202	Hard disk20G	1
6	3003-20-35008	Rubber washer	6
11	M04-000301---	Stainless steel hexagon nut GB/T6170-2000 M3 polished	4
12	M90-000170---	FAN 12V 8.0CFM 32dBA 40*40*10mm	1
13	6800-20-50147	CIS fan support rack	1
7	6800-20-50397	Vent baffle 2	1

6.5.6 Others

P/N	Description	Qty
CIS assembly －6800-30-50474		
0000-10-11201	Memory bank	1
0000-10-11202	Hard disk	1
3003-20-35008	Rubber washe	6
0000-10-10755	Control board	1
6800-20-50304	Ethernet cable for DVI interface board	1
6800-20-50333	CIS board flat cable	1
6800-20-50387	CIS hard disk insulating plate	1
6800-20-50550	CIS hard disk support	1
6800-20-50690	Hard disk signal wir	1
6800-20-50692	Backlight board thermo-conductive insulating plate	1
6800-20-50700	Heat conducting block for CIS hard disk	1
6800-30-50112	CIS mother board	1
M04-004012---	Crosshead screw	3
M04-051045-00	Stainless steel crosshead screw	4
M05-100R29-02	Lithium button cell battery, 3V 220mAh D20*3.2	1
TR6F recorder－TR6F-30-67306		
0000-10-11079	Thermal print head	1
9201-20-36007	Grounding wire	1
M04-051003---	Crosshead tapping screw PT2X6	6
TR6F-20-67300	Recorder chamber	1
TR6F-20-67301	Recorder door	1
TR6F-20-67302	Spanner	1
TR6F-20-67303	Internal adapter	1
TR6F-20-67304	Back spring	1
TR6F-20-67305	Silicon button	1
TR6F-20-67307	Overlay	1
TR6F-20-67314	Cable from recorder drive board to recorder button board	1
TR6F-30-67308	Recorder button board	1
TR6F-30-67310	Recorder drive board	1

6.6 SMR Assembly

6.6.1 SMR Assembly

Exploded View



Parts List

SN	P/N	Description	Qty
1	6800-20-50249-51	Inside assembly	1
2	M6G-020015---	Hose	1
3	6800-30-51154	Interface board	1
4	6800-20-50250	Rear housing	1
5	6800-20-50254	Side plate fixture	2
6	M04-004015---	Crosshead screw M3×8	4
7	6800-20-50308	Screw cap 2	4
8	6800-20-50253	Side plate, right	1
9	6800-20-50262	Rubber washer	1
10	6800-20-50251	Handle	1
11	6800-20-50252	Side plate, left	1
12	6800-20-50263	Cushion	4

6.6.2 SMR Inside Assembly

Exploded View



Parts List

SN	P/N	Description	Qty
1	6800-20-50519	Insulating plate	1
2	9201-20-36010	Torsion spring washer	2
3	6800-20-50256	Light conducting pole	1
4	6800-20-50261	Contact spring	16
5	6800-21-51100	Contact screw	16
6	6800-20-50279	Infrared light filter	8
7	6800-20-50249	SMR inside assembly	1
8	6800-20-50388	Leaf	16
9	M04-021024---	Washer	16
10	M04-011002---	Hex nut assy	16
11	6800-20-50257-51	Reinforced section steel	1
12	6800-30-51153	8-slot module rack communication board	1
13	6800-30-50075	Nios II module	1
14	M04-004012---	Crosshead screw M3×6	8

6.7 Parameter Modules

6.7.1 MPM Module

Exploded View



Parts List

SN	P/N	Description	Qty
1	6800-20-50278	Spanner	2
2	6800-20-50293	Screw	2
3	6800-20-50279	Infrared light filter	1
4	M04-000104---	Spring washer	2
5	6800-20-50273	Rear cover	1
6	6800-20-50339	Waterproof seal 02	1
7	6800-30-50684	MPM module infrared communication backboard	1

8	6800-20-50275	Right cover	1
9	M51A-30-80924	NIBP&MPM analog module	1
10	M51A-30-80870	MPM front panel assembly	1
11	M04-005005---	Crosshead screw M3×6	2
12	6800-20-50274	Left cover	1
13	M04-011002---	M3 nut with spring washer	2
14	6800-20-50292	Fixer	2
15	6800-20-50870	Cushion	2

6.7.2 C.O. Module

Exploded View



Parts List

SN	P/N	Description	Qty
1	6800-20-50264	Front panel	1
2	6800-20-50284	Silicon button	1
3	6800-20-50282	Silicon button fixture	1
4	6800-30-50114	Button board	1
5	M04-003105---	Tapping screw PT3×8	1
6	6800-20-50269	Left cover	1
7	6800-20-50338	Waterproof seal 01	1
8	6800-20-50268	Rear cover	1
9	6800-20-50279	Infrared light filter	1
10	6800-20-50293	Screw	2
11	M04-000104---	Spring washer	2
12	6800-20-50278	Spanner	1
13	6800-20-50292	Fixer	2
14	6800-30-50798	CO module infrared communication board	1
15	M04-011002---	M3 nut with spring washer	2
16	M03A-30-26055	CO board	1
17	6800-20-50270	Left cover	1
18	6800-20-50101	C.O. connector	1
19	M04-005005---	Crosshead screw M3×6	1
20	6800-20-50394	Snap close	1
21	6800-20-50286	Locking device	1

6.7.3 RM Module

Exploded View



Parts List

SN	P/N	Description	Qty
1	6800-20-50264	Front panel	1
2	6800-20-50284	Silicon button	1
3	6800-20-50282	Silicon button fixture	1
4	6800-30-50114	Button board	1
5	M04-003105---	Tapping screw PT3×8	1
6	6800-20-50269	Left cover	1
7	6800-20-50338	Waterproof seal 01	1
8	6800-20-50268	rear cover	1
9	6800-20-50293	screw	2
10	M04-000104---	Spring washer	2
11	6800-20-50279	Infrared light filter	1

12	6800-20-50278	Spanner	1
13	6800-30-50812	RM module infrared communication board	1
14	M04A-30-58452	RM main unit	1
15	6800-20-50270	right cover	1
16	M04-005005---	Crosshead screw M3×6	1
17	6800-30-50152	Infrared detecting board	1
18	6800-20-50394	Snap close	1
19	6800-20-50286	Locking device	1
20	6800-30-50489	Connector assembly	1

6.7.4 ICG Module

Exploded View



Parts List

SN	P/N	Description	Qty
1	0010-10-42826	Signal wire	1
2	6800-20-50266	Front panel	1
3	6800-20-50284	Silicon button	1
4	6800-20-50282	Silicon button fixture	1
5	6800-30-50114	Button board	1
6	M04-003105---	Tapping screw PT3×8	1
7	0010-10-42827	ICG module PCB assembly	1
8	6800-20-50269	Left cover	1
9	6800-30-50094	Non isolating power board	1
10	6800-20-50338	Waterproof seal 01	1
11	6800-20-50268	Rear cover	1
12	6800-20-50279	Infrared light filter	1
13	M04-000104---	Spring washer	2
14	6800-20-50293	Screw	2
15	6800-20-50278	Spanner	1
16	6800-20-50292	Fixer	2
17	6800-30-50802	ICG module infrared communication board	1
18	M04-011002---	M3 NUT WITH SPRING WASHER	2
19	6800-20-50270	Right cover	1
20	M04-005005---	Crosshead screw M3×6	1
21	6800-20-50394	Snap close	1
22	6800-20-50286	Locking device	1
23	M39-059004---	Wire fixture	1

6.7.5 AG Module

Exploded View



Parts List

SN	P/N	Description	Qty
1	6800-30-50539	Front panel assembly (with BIS)	1
2	M04-005005---	Crosshead screw M3×6	2
3	9200-21-10605	AG module box	1
4	6800-20-50290	Aluminium support rack	1
5	9200-10-10529	AG module	1
6	M39-000046---	TIE MOUNT	3
7	M04-011002---	M3 nut with spring washer	4
8	6800-30-50810	AG module Infrared communication back board	2
9	6800-20-50292	Fixer	4
10	6800-20-50278	spanner	2
11	6800-20-50293	screw	4
12	M04-000104---	Spring washer	4
13	6800-20-50279	Infrared light filter	2

14	6800-20-50277	Rear cover	1
15	6800-20-50340	Waterproof seal 03	1
16	M04-006512---	Crosshead screw M4×6	2
17	6800-20-50291	O_2 module mount	1
18	M04-004012---	Crosshead screw M3×6	8
19	9200-10-10531	O_2 module	1
20	6800-20-50281	Infrared communication back board support rack	1
21	6800-30-50092	Isolating power board	1
22	M04-000405---	Crosshead screw M3×8	8
23	6800-20-50699	Insulating pad	4

6.7.6 BIS Module

Exploded View



Parts List

SN	P/N	Description	Qty
1	6800-20-50264-51	(BIS) Front panel	1
2	6800-20-50104	BIS connector	1
3	6800-20-50284	Silicon button	1
4	6800-20-50282	Silicon button fixture	1
5	6800-30-50114	Button board	1
6	M04-003105---	Tapping screw PT3×8	1
7	6800-20-50269	Left cover	1
8	6800-30-50092	Isolating power board	1
9	6800-30-50800	BISX module infrared communication board	1
10	6800-20-50286	Locking device	1
11	6800-20-50338	Waterproof seal 01	1
12	6800-20-50268	Rear cover	1
13	6800-20-50279	Infrared light filter	1
14	6800-20-50293	Screw	2
15	M04-000104---	Spring washer	2
16	6800-20-50278	Spanner	1
17	6800-20-50292	Fixer	2
18	M04-005005---	Crosshead screw M3×6	1
19	6800-20-50394	Snap close	1
20	M04-011002---	M3 nut with spring washer	2
21	6800-20-50270	Right cover	1

6.7.7 IBP Module

Exploded View



Parts List

SN	P/N	Description	Qty
1	6800-20-50265	Front panel(IBP)	1
2	6800-20-50286	Locking device	1
3	M04-005005---	Crosshead screw M3×6	1
4	6800-20-50100	IBP connector	2
5	M03a-30-26058	IBP board	1
6	6800-20-50269	Left cover	1
7	M04-011002---	M3 nut with spring washer	2
8	6800-20-50292	Fixer	2
9	6800-20-50394	Snap close	1
10	6800-20-50278	Spanner	1
11	6800-20-50293	Screw	2
12	M04-000104---	Spring washer	2
13	6800-20-50279	Infrared light filter	1
14	6800-20-50268	Rear cover	1

15	6800-20-50338	Waterproof seal 01	1
16	6800-30-50796	IBP module Infrared communication back board	1
17	6800-20-50270	Right cover	1
18	M04-003105---	Tapping screw PT3×8	1
19	6800-30-50114	Button board	1
20	6800-20-50282	Silicon button fixture	1
21	6800-20-50284	Silicon button	1

6.7.8 Mindray CO2 Module

Exploded View



Parts List

SN	P/N	Description	Qty
1	6800-20-50275	Right cover	1
2	6800-30-50538	Mindray CO_2 module front panel	1
3	M02B-30-64513	CO2 module main unit	1
4	M04-005005---	Crosshead screw M3×6	2
5	6800-20-50274	Left cover	1
6	6800-30-50804	Mindray CO2 module infrared communication board	1
7	6800-20-50339	Waterproof seal 02	1
8	M04-011002---	M3 nut with spring washer	2
9	6800-20-50292	Fixer	2
10	6800-20-50278	Spanner	2
11	6800-20-50273	Rear cover	1
12	M04-000104---	Spring washer	2
13	6800-20-50279	Infrared light filter	1
14	6800-20-50293	Screw	2

6.7.9 Oridion CO2 Module

Exploded View



Parts List

SN	P/N	Description	Qty
1	6200-20-11614	Outlet	1
2	9201-20-35914	Connector cover	1
3	6800-20-50267	Front panel	1
4	M04-000501---	Stainless steel nut	1
5	6800-20-50284	Silicon button	1
6	6800-20-50282	Silicon button fixture	1
7	6800-30-50114	Button board	1
8	0010-10-42559	Microstream CO2 module	1
9	6800-30-50094	Non isolating power board	1
10	6800-20-50269	Left cover	1
11	6800-20-50268	Rear cover	1
12	6800-20-50279	Infrared light filter	1
13	6800-20-50293	Screw	2
14	M04-000104---	Spring washer	2

15	6800-20-50278	Spanner	1
16	6800-20-50338	Waterproof seal 01	1
17	6800-20-50292	Fixer	2
18	6800-30-50806	Oridion CO2 module infrared communication board	1
19	M04-011002---	M3 nut with spring washer	2
20	6800-20-50270	Right cover	1
21	9201-30-35959	Connector	1
22	M04-005005---	Crosshead screw M3×6	1
23	M04-003105---	Tapping screw PT3×8	2
24	6800-20-50286	Locking device	1
25	9201-20-36010	Torsion spring washer	1
26	6800-20-50394	Snap close	1
27	9201-20-35961	Torsion spring	1

6.8 Wireless AP

Exploded View



Part List

SN	P/N	Description	Qty
1	9201-20-36008-51 or 9201-20-35916-51(configured with Dlink AP)	(ASUS)Wireless network card fixed board Wireless network card fixed board (DLINK)	1
2	M04-004015---or M04-051147---(configured with Dlink AP)	Cross recessed small pan head screw with washer M3X8 Crosss recessed countersunk head screw M3X8	2
3	9201-20-36107	Wireless network card locking plate	1
4	9201-20-36109	Wireless network card spacer	2
5	9201-20-36009-51 or 9201-20-35917-51(configured with Dlink AP)	Wireless network card installation board (ASUS) Wireless network card installation board (DLINK)	1
6	9201-20-36110	Overlay of wireless network card fixed board	1
7	0000-10-10777 or 0000-10-11296	Wireless AP 54Mbps mini-type (ASUS) Wireless AP 2.4GHz 54Mbps(DLINK)	1

6.9 Replaceable Parts

To replace the parts, please refer to ***5 Repair and Disassembly*** and the exploded views above.

NOTE

- **Here we list most replaceable parts. If you need more parts, please contact our Customer Service Department.**

6.9.1 Main Unit

P/N	Description	Qty
Rear housing assembly		
6800-20-50194	Recorder cover	1
TR6F-30-67306	TR6F recorder	1
6800-20-50301	Recorder to main board cable	1
6800-20-50197-51	Mains container	1
6800-20-50298	Cable, power board to battery interface board	1
6800-20-50155	Cable, main unit to infrared communication board	1
6800-30-50071	3-slot module rack communication board	1
6800-30-50073	MPM module rack communication board	1
6800-30-50075	NiosII module	1
6800-30-50124	Integral module rack flexible cable	1
0000-10-11063	CF card	1
6800-20-50198-51	Key pad	1
6800-20-50184-51	Rear housing	1
6800-20-50672	Cable, main board to DIV interface board	1
6800-20-50673	Cable, main board to USB interface board	1
6800-30-50532	Main board assembly	1
6800-30-50401	WLAN&CF card flexible cable	1
6800-30-50472	CF/WLAN assembly	1
6800-30-50476	Interface board assembly	1
6800-30-50509	Fan assembly	1
6800-20-50681	Speaker & cable	1

6800-30-50474	CIS assembly	1
Front housing assembly		
0000-10-11072	Backlight board	1
6800-20-50177-52	Front bezel	1
0000-10-10754 or 0000-10-11069（optional ）	LCD screen	1
0000-10-11071 or 0000-10-10748（optional）	Touchscreen	1
6800-30-50082 or 6800-30-51095（optional）	Touchscreen control board (only used together with optional touchscreen0000-10-10748)	1
6800-30-50888	Button board (1280x1024-1+Touch) (without shank)	1
6800-30-50086	Alarm LED board	1
6800-30-51132	Encoder assembly(without shank)	1
Base assembly		
6800-20-50181-51	Bottom cover	1
6800-30-50108	Battery interface board	1
6800-30-50088	Power switch board	1

6.9.2 SMR

P/N	Description	Qty
6800-30-50075	NiosII module	1
6800-30-50078	SMR communication board	1
6800-30-50080	SMR interface board	1
6800-30-50667	SMR indicating lamp assembly	1
6800-20-50249-51	SMR inside assembly (BeneView)	1
6800-20-50250-51	SMR rear housing (BeneView)	1
6800-20-50251	Handle	1

6.9.3 Parameter Modules

P/N	Description	Qty
6800-30-50486	BIS Module	1
6800-30-50488	RM module	1
6800-30-50487	CAPNOSTAT CO_2 module	1
6800-30-50559	ORIDION CO_2 module	1
6800-30-50500	Mindray CO_2 module	1
6800-30-50491	ICG module	1
6800-30-50501	AG module (with O_2/BIS)	1
6800-30-50502	AG module (with O_2)	1
6800-30-50503	AG module (without O_2/BIS)	1
6800-30-50504	AG module (without O_2)	1
6800-30-50505	Manual AG module (with O_2/BIS)	1
6800-30-50506	Manual AG module (with O_2)	1
6800-30-50507	Manual AG module (without O_2/BIS)	1
6800-30-50508	Manual AG module (without O_2)	1
6800-30-50484	C.O. module	1
6800-30-50485	IBP module	1
M51A-30-80873	MPM-1 module(MINDRAYSpo2/3/5 lead)	1
M51A-30-80874	MPM-2 module(MasimoSpo2/3/5 lead)	1
M51A-30-80877	MPM-3 module(NellcorSpo2/3/5 lead)	1

M51A-30-80880	MPM-4 module (MINDRAYSpo2/12 lead)	1
M51A-30-80881	MPM-5 module(MasimoSpo2/12 lead)	1
M51A-30-80884	MPM-6 module(NellcorSpo2/12 lead)	1
M51A-30-80900	MPM-7 module main unit(MindraySpo2/3/5 lead/without IBP)	1
M51A-30-80901	MPM-8 module(Masi Spo2/3/5 lead/without IBP)	1
M51A-30-80904	MPM-9 module(Nell Spo2/3/5 lead/without IBP)	1
M51A-30-80907	MPM-10 module main unit (MindraySpo2/12 lead/without IBP)	1
M51A-30-80908	MPM-11 module (Masi Spo2/12 lead /without IBP)	1
M51A-30-80911	MPM-12 module (Nell Spo2/12 lead/without IBP)	1

6.9.4 Cables

P/N	Description	Remarks
Main unit		
6800-20-50156	Alarm LED board cable	
6800-20-50157	Touchscreen control board cable	
6800-20-50159	Power switch & LED board cable	
6800-20-50298	Cable from power board to battery interface board	
6800-20-50301	Cable from recorder to main board	
6800-20-50304	DVI interface board signal wire	DVI interface board to CIS assembly
6800-20-50305	power board DC output cable	
6800-20-50333	CIS mother board flat wire	Mother board to CIS mother board
6800-20-50334	17” LCD flat wire	Mother board to LCD
6800-21-50337	AC input filter cable	
6800-20-50513	Main unit Fan cable	
6800-20-50672	Cable from mother board to DVI interface board	
6800-20-50673	Cable from mother board to USB interface board	
6800-20-50681	Speaker & cables	
6800-30-50124	Integral module rack flexible cable,	Mother board to integral module rack
6800-30-50126	Button board flexible cable	Mother board to button board

6800-20-50401	WLAN card flexible cable	Mother board to CF&WLAN assembly
SMR and parameter modules		
6800-20-50155	Main unit infrared communication board cable (integral module rack)	
6800-20-50160	Button board cable	
6800-20-50161	Cable from infrared communication board to RS232 connector	
6800-20-50162	Infrared communication board TTL cable	
6800-20-50164	Module fan & cable	
6800-20-50167	SMR LED cable	
6800-20-50170	Cable from inverter to button board	from inverter to button board
6800-20-50306	AG cable	
6800-20-50309	ICG communication cable	
6800-21-50310	BIS interface cable	
6800-21-50311	CO interface cable	
6800-21-50312	IBP interface cable	
6800-21-50313	Mainstream CO_2 interface cable	
6800-20-50315	ICG power cord	
6800-20-50316	RM infrared detection board cable	
6800-20-50319	Cable from SMR to main unit	
6800-20-50662	gas pump cable	NIBP pump to parameter board
6800-20-50663	Fast-release valve cable	Fast-release valve to parameter board
6800-20-50664	Slow-release valve cable	Slow-release valve to parameter board
6800-20-50674	Cable from ICG module to infrared communication board	
6800-20-50683	Mindray CO_2 infrared communication cable	
6800-20-50690	CIS Hard disk cable	
6800-30-50128	Mindray SpO_2 flexible cable kit	
6800-30-50132	Nellcor SpO_2 flexible cable kit	
6800-30-50130	Masimo SpO_2 flexible cable kit	

FOR YOUR NOTES